



Republic of the Philippine
TARLAC STATE UNIVERSITY
Romulo Blvd., San Vicente, Tarlac City
Tel. No.: (045) 982 4630
Website: www.tsu.edu.ph

Bidding Documents

(This Bidding Documents is in conformance with the Sixth Edition of the Philippine Bidding Documents for the Procurement of Goods)

For the Project

Increase in Carrying Capacity of Nursing and Allied Health Programs

**With an Approved Budget for the Contract (ABC) of
Fifty Million Pesos (₱ 50,000,000.00)**

**Invitation to Bid No. Goods 2023-050
PhilGEPS Reference No.: 10138727**

**July 2020
6th Edition**

TABLE OF CONTENTS

Glossary of Acronyms, Terms, and Abbreviations	1
Section I. Invitation to Bid	4
Section II. Instructions to Bidders	6
1. Scope of Bid	6
2. Funding Information	6
3. Bidding Requirements	6
4. Corrupt, Fraudulent, Collusive, Coercive, and Obstructive Practices	6
5. Eligible Bidders	6
6. Origin of Goods	7
7. Subcontracts	7
8. Pre-Bid Conference	7
9. Clarification and Amendment of Bidding Documents	7
10. Documents Comprising the Bid: Eligibility and Technical Components	7
11. Documents Comprising the Bid: Financial Component	7
12. Bid Prices	8
13. Bid and Payment Currencies	8
14. Bid Security	8
15. Sealing and Marking of Bids	8
16. Deadline for Submission of Bids	9
17. Opening and Preliminary Examination of Bids	9
18. Domestic Preference	9
19. Detailed Evaluation and Comparison of Bids	9
20. Post Qualification	10
21. Signing of the Contract	10
Section III. Bid Data Sheet	11
Section IV. General Conditions of Contract	12
1. Scope of Contract	12
2. Advance Payment and Terms of Payment	12
3. Performance Security	12
4. Inspection and Tests	12
5. Warranty	12
6. Liability of the Supplier	13
Section V. Special Conditions of Contract	14
Section VI. Schedule of Requirements	16
Section VII. Technical Specifications	17
Section VIII. Checklist of Technical and Financial Documents	19
Section IX. Bidding Forms	20

Glossary of Acronyms, Terms, and Abbreviations

ABC – Approved Budget for the Contract.

BAC – Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR – Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF – Cost Insurance and Freight.

CIP – Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP – Refers to the quoted price of the Goods, which means “delivered duty paid.”

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA – “Free Carrier” shipping point.

FOB – “Free on Board” shipping point.

Foreign-funded Procurement or Foreign-Assisted Project– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as “Call-Offs,” are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term “related” or “analogous services” shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB – Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs – Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Section I. Invitation to Bid



Republic of the Philippine
TARLAC STATE UNIVERSITY
Romulo Blvd., San Vicente, Tarlac City
Tel. No.: (045) 982 4630
Website: www.tsu.edu.ph

INVITATION TO BID

For the Project

Increase in Carrying Capacity of Nursing and Allied Health Programs

Invitation to Bid No. Goods 2023-050

1. The Tarlac State University, through General Appropriation Act intends to apply the sum of **Fifty Million Pesos (₱ 50,000,000.00)** to payments under the contracts for the project: **Increase in Carrying Capacity of Nursing and Allied Health Programs**, comprising of the following lots with their corresponding ABCs:

Lot No.	Description	ABC, ₱
1	Manikin	45,695,000.00
2	Hospital Equipment	4,305,000.00

Bids received in excess of the ABC for shall be automatically rejected at bid opening.

2. The Tarlac State University now invites bids for the project **Increase in Carrying Capacity of Nursing and Allied Health Programs**. Delivery of the Goods is required within **ninety (90)** calendar days from the receipt of the Notice to Proceed. Bidders must have completed, within the last five (5) years prior to the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instructions to Bidders.
3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary “pass/fail” criterion as specified in the 2016 Revised Implementing Rules and Regulations (IRR) of Republic Act (RA) 9184, otherwise known as the “Government Procurement Reform Act”.

Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA 5183.

4. Interested bidders may obtain further information from **Tarlac State University** and inspect the Bidding Documents at the address given below during Tuesday to Friday from 7:00 A.M. to 6:00 P.M:

BAC Secretariat
Motorpool and Administration Building
Tarlac State University
Romulo Blvd., San Vicente, Tarlac City
Tel. No. (045) 606-8142 / 0998 846 0206
Email: bacsec@tsu.edu.ph

5. A complete set of Bidding Documents may be acquired by interested Bidders from **September 18, 2023 to October 10, 2023** from the aforementioned address upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, as indicated below:

Lot No.	Bidding Documents Fee, ₱
1	25,000.00
2	5,000.00

It may also be downloaded free of charge from the website of the Philippine Government Electronic Procurement System (PhilGEPS) and the website of the Procuring Entity, provided that Bidders shall pay the applicable fee for the Bidding Documents not later than the submission of their bids.

6. The Tarlac State University will hold a Pre-Bid Conference on **September 26, 2023 (2:00 P.M.)** at the Business Center Audio-Visual Room, 2nd Floor, Business Center Bldg., Tarlac State University, Romulo Blvd., San Vicente, Tarlac City, which shall be open to prospective bidders.
7. Bids must be duly received by the BAC Secretariat at the address below on or before **October 10, 2023 (2:00 P.M.)**.

BAC Secretariat
Motorpool and Administration Building
Tarlac State University
Romulo Blvd., San Vicente, Tarlac City
Tel. No. (045) 606-8142 / 0998 846 0206
Email: bacsec@tsu.edu.ph

8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 14.
9. Bid opening shall be on **October 10, 2023 at 2:00 P.M.**, at the Business Center Audio-Visual Room, 2nd Floor, Business Center Bldg., Tarlac State University, Romulo Blvd., San Vicente, Tarlac City. Bids will be opened in the presence of the bidders' representatives who choose to attend at the address below.
10. The Tarlac State University reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 Revised IRR of RA 9184, without thereby incurring any liability to the affected bidder or bidders.
11. For further information, please refer to:

Ms. Jhenna Micah A. Manankil/ Mr. Joshua Jonathan S. Jacinto
BAC Secretariat
Motorpool and Administration Building
Tarlac State University
Romulo Blvd., San Vicente, Tarlac City
Tel. No. (045) 606-8142 / 0998 846 0206
Email: bacsec@tsu.edu.ph



DR. MURPHY P. MOHAMMED
BAC Chairperson

Section II. Instructions to Bidders

1. Scope of Bid

- 1.1. The Tarlac State University wishes to receive Bids for the project: **Increase in Carrying Capacity of Nursing and Allied Health Programs**, with identification number **Invitation to Bid No. Goods 2023-050**.
- 1.2. The procurement project (referred to herein as “Project”) is composed of two (2) lots the details of which are described in Section VII. Technical Specifications.

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for **2023** in the amount of **₱ 50,000,000.00**
- 2.2. The source of funding is the **General Appropriation Act**.

3. Bidding Requirements

- 3.1. The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.
- 3.2. Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or IB by the BAC through the issuance of a supplemental or bid bulletin.
- 3.3. The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex “I” of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership limited to those allowed under the rules may participate in this Project.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder must have a SLCC that is similar to the Project, as described in the **BDS**, with a value, adjusted to current prices using the PSA’s CPI, that is at least equivalent to fifty percent (50 %) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that subcontracting is not allowed

8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and at the address indicated in paragraph 6 of the **IB**.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII. Checklist of Technical and Financial Documents**.

10.2. The Bidder's **SLCC** as indicated in **ITB** Clause 5.3 should have been completed **within five (5) years** prior to the deadline for the submission and receipt of bids.

10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII. Checklist of Technical and Financial Documents**.

11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.

11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.

11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:

- a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.
- b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in **Section VII. Technical Specifications**.

13. Bid and Payment Currencies

13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.

13.2. Payment of the contract price shall be made in Philippine Pesos.

14. Bid Security

14.1. The Bidder shall submit a Bid Securing Declaration or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.

14.2. The Bid and bid security shall be valid until **February 07, 2024**. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

15.1. Each Bidder shall submit one copy of the first and second components of its Bid.

15.2. The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

15.3. If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be

digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the **IB**.

17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.

19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.

19.3. The descriptions of the lots or items are indicated in **Section VII. Technical Specifications**, and the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.

19.4. The Project shall be awarded as one project having several lots that shall be awarded as separate contracts.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

- 20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

21. Signing of the Contract

- 21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Section III. Bid Data Sheet

ITB Clause	
5.3	For this purpose, contracts similar to the Project shall be the supply and delivery of similar goods comprising the Project.
7.1	No further instructions.
12	The price of the Goods shall be quoted DDP Tarlac City, Philippines or the applicable International Commercial Terms (INCOTERMS) for this Project.
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <ul style="list-style-type: none"> a. The amount of not less than two percent (2 %) of ABC, if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than five percent (5 %) of ABC if bid security is in Surety Bond.
19.3	The ABC for the project is ₱ 50,000,000.00
20.2	No further requirements.
21.2	No further requirements.

Section IV. General Conditions of Contract

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

2.1. Advance payment of the contract amount is provided under Annex “D” of the revised 2016 IRR of RA No. 9184.

2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the **SCC, Section IV. Technical Specifications** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.

6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the

defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

GCC Clause	
1	<p>Delivery and Documents –</p> <p>For purposes of the Contract, “EXW,” “FOB,” “FCA,” “CIF,” “CIP,” “DDP” and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:</p> <p>The delivery terms applicable to this Contract are delivered to Tarlac State University, Romulo Blvd., San Vicente, Tarlac City. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their destination.”</p> <p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI. Schedule of Requirements.</p>
	<p>Packaging –</p> <p>The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods’ final destination and the absence of heavy handling facilities at all points in transit.</p> <p>The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.</p> <p>The outer packaging must be clearly marked on at least four (4) sides as follows:</p> <p>Name of the Procuring Entity Name of the Supplier Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications</p> <p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p>
	<p>Transportation –</p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p>

	<p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p>
	<p>Intellectual Property Rights –</p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
4	<p>The inspections and tests that will be conducted are: visual and sensory inspection and test.</p>

Section VI. Schedule of Requirements

The delivery date for the Goods covered by the Contract shall be within ninety (90) calendar days upon receipt of the Notice to Proceed.

Section VII. Technical Specifications

Item No.	Description
LOT 1 – MANIKIN	
1.	<p>PEDIATRIC SIMULATOR (HIGH FEDELITY)</p> <p>General</p> <ul style="list-style-type: none"> • Height: 44 inches • Tetherless and wireless; fully responsive during transport • The internal rechargeable battery provides hours of tetherless operation • Smooth and supple full-body skin with seamless trunk and limb joints • Realistic joint articulation: neck, shoulder, elbow, hip, and knee • Palpable bony landmarks • Forearm pronation and supination • Supports common patient positions including Fowler’s, supine, and sitting • Male/female patient conversion • Tablet PC preloaded with UNI® included • Includes 10 preprogrammed SLEs and facilitator’s guidebook <p>Neurological</p> <ul style="list-style-type: none"> • Active robotics simulate lifelike facial expressions including: anger, transient pain, ongoing pain, amazement, Quizzical, crying, yawning <p>Preprogrammed emotional states automatically express associated verbal and non-verbal cues without manual input: Worried, Anxious, Lethargic, Distracted</p> <ul style="list-style-type: none"> • Abnormal eye and eyelid movements: crosseyed, nystagmus, eyelid twitching, eyelid droop, Programmable crying/ tears release real fluid, Wireless streaming voice: listen to participants respond in real-time, Real-time voice modulation effects, Automatic jaw movement synchronized with speech, Automatic jaw movement synchronized with speech, Seizures with selectable intensity levels, with prerecorded speech responses • Real CO2 exhalation: supports etCO2 monitoring using real sensors and monitoring devices, Selectable normal and abnormal sounds: upper right, front and back; upper left, front and back; lower right, back; and lower left, back, uses “Real” mechanical ventilation support, Real-time ventilation feedback, Visible chests rise during BVM ventilation, Chest tube insertion: left midaxillary hemothorax site features palpable bony landmarks, realistic skin for cutting and suturing, tactile pleural pop, and fluid drain, Needle decompression site features realistic tactile feedback and audible hiss, Needle decompression and chest tube insertion <p>Cardiac</p> <ul style="list-style-type: none"> • Includes comprehensive library of ECG rhythms with customizable beat variations, Independent normal/abnormal heart sounds at aortic, pulmonic, and mitral sites, Supports ECG monitoring using real devices Supports ECG-derived respiration monitoring (EDR), Ecpr™ Real-time quality feedback and reporting Defibrillate, cardiovert and pace using “real devices” and energy Circulatory, Visible cyanosis, redness, pallor, and jaundice, Supports capillary refill time testing above the right knee; test detection and logging, Palpable pulses: bilateral carotid, brachial, radial, femoral, and pedal, Blood pressure-dependent pulses, Supports blood pressure monitoring using a “real NIBP” cuff and monitor, SpO2 monitoring using real devices Vascular Access, Bilateral forearm IV access supports sampling and continuous infusion, Intraosseous infusion site at right proximal tibia, Real glucose test readings via finger-stick <p>Gastrointestinal, Patent esophagus, Gastric distension during excessive PPV, Bowel sounds in four quadrants, Interchangeable male/female genitalia, Supports urinary catheterization with fluid return, Programmable urinary output</p> <p>Airway</p> <ul style="list-style-type: none"> • Anatomically accurate oral cavity and airway, Supports nasotracheal/ orotracheal

	<p>intubation with standard instruments including endotracheal tubes and supraglottic airway devices, Tracheal intubation detection, Head tilt, chin lift, jaw thrust, Supports esophageal intubation, NG/OG tube placement, Supports bag-valve-mask ventilation, Realistic surgical trachea permits tracheostomy, cricothyrotomy, and retrograde intubation, Programmable difficult airway: laryngospasm and tongue edema, Selectable normal and abnormal upper airway sounds</p> <p>Breathing</p> <ul style="list-style-type: none"> •Spontaneous breathing and selectable normal and abnormal respiratory patterns •Variable respiratory rates and inspiratory/expiratory ratios •Programmable unilateral chest rise and fall •Unilateral chest rise with right mainstem intubation <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . Training 2-3 days.</p>
2.	<p>Maternal and Neonatal Birthing Simulator (ULTRA HIGH FIDELITY MANIKIN)</p> <p>GENERAL SPECIFICATION</p> <ul style="list-style-type: none"> •Tetherless and wireless; fully responsive during transport, Wireless control at distances up to 300 ft. Internal rechargeable battery provides up to 10 hrs. of tetherless operation, Smooth and supple full-body skin with seamless articulating joints, Lifelike joint articulation: neck, shoulder, elbow, wrist, hip, knee, and ankle, Internal high-capacity fluid reservoirs, Maternal® Fetus-Newborn wireless link capability, Includes touchscreen tablet PC preloaded with software, Includes touchscreen virtual patient monitor and CTG <p>TECHNICAL DATA SHEET</p> <p>Processor : 12th Gen Intel® Core™ i5-12500 (18 MB cache, 6 cores, 12 threads, 3.00 GHz to 4.60 GHz Turbo, 65 W) Operating System : Windows 11 Pro Graphics Card : Intel® Graphics Monitor : OptiPlex All-in-One Non-Touch Panel Memory * : 8 GB, 1 x 8 GB, DDR4 Hard Drive : 256 GB, M.2 2230, PCIe NVMe, SSD, Class 35 Keyboard : Wireless Keyboard and Mouse Ports on the Side : 1 USB 3.2 Gen 2x1 Type-C® port 1 USB 3.2 Gen 1 port with PowerShare 1 Universal audio port Rear : 2 USB 3.2 Gen 2 ports 2 USB 3.2 Gen 1 ports with Smart Power On 1 RJ45 Ethernet port 1 Line-out audio port 1 DisplayPort++ 1.4a/HDCP 2.3 port 1 HDMI-IN—HDMI 1.4a port (with UMA configuration-65 W) 1 HDMI-OUT—HDMI 2.0 port (with UMA configuration-65 W) Slots : 1 SD-card slot (side) 1 M.2 2230 slot for Wi-Fi and Bluetooth card 1 M.2 2230/2280 slot for SSD (35 W) 2 M.2 2230/2280 slot for SSD (65 W)</p> <p>NEUROLOGICAL</p> <ul style="list-style-type: none"> •Programmable blinking rate, pupil response, and bilateral and unilateral eye movement, Interactive eyes can follow a moving object, Seizures with selectable

intensity levels, Wireless streaming voice: be the voice of Victoria and listen to participants' responses via a wireless headset, Record and playback vocal responses in any language

AIRWAY: Head tilt/chin lift/jaw thrust, Oral or nasal intubation, Difficult airway: laryngospasm, tongue edema, and pharyngeal swelling.

BREATHING: Spontaneous breathing; selectable respiratory patterns and lung sounds, Programmable respiratory rates and inspiratory/expiratory ratios Realistic chest rise during assisted ventilation, Ventilation performance, real-time monitoring, reporting, and logging,

CARDIAC: Real-time compression and ventilation performance, Effective compressions generate palpable pulses and ECG artifacts. Real-time 4-Lead electrocardiogram monitoring using real ECG devices, Defibrillation, cardioversion, and pacing using real energy, Realistic heart sounds, Programmable heart rate synchronized with ECG and pulses, Virtual 12-lead dynamic ECG rhythms

CIRCULATION

•Bilateral palpable carotid, radial, and brachial pulses are synchronized with heart rate and blood pressure, Monitor oxygen saturation using real oximeters, Measure blood pressure using a real manual or automatic blood pressure cuff, Bilateral venous access, Automatic drug recognition detects medication type, dose, and rate injected into the lower right arm, SubQ and intramuscular injection sites for placement exercises

OBSTETRICS

•Maternal-fetal physiologic link, Supports Leopold's Maneuvers and external cephalic version, Supports fetal monitoring using real devices, Precision Delivery System: computer-controlled fetal descent and cardinal movements, Programmable normal, breech, shoulder dystocia, instrument assisted delivery, and C-section, Software activated self-lubricating birth canal, Realistic placenta with detachable fragments and umbilical cord

OBSTETRICS

•Epidural placement and needle detection; palpable anatomical landmarks and skin layers, Intrapartum bleeding, Force sensor reports traction and torsion applied to the fetus in real-time, McRoberts maneuver, suprapubic pressure, Zavanelli, Woods' Screw, and "hands and knees" position detection, Births full-term neonate of realistic size and weight, Active neonate: programmable heart and respiratory sounds, crying, cyanosis, and movement, Replaceable episiotomy inserts support suturing, Multi-layer fourth degree episiotomy with hemorrhaging vaginal sidewall tears and cervical lacerations, Programmable uterine firmness and hemorrhaging supports sutures and balloon tamponade, Programmable hemorrhage flow control

GASTROINTESTINAL

•Selectable bowel sounds

•Internal fluid bladder with urethra for Foley catheterization exercises

•Rectum with suppository placement sensor

Package inclusions:

•Maternal Birthing Simulator, Active birthing baby, Tablet PC with rugged case, Simulator Control Software with Lifetime License, Vitals Patient Monitor, (9) Labor & Delivery Simulation Learning Experience, Automatic Mode control license, Integrated automatic drug recognition, 20 Programmable drug recognition syringes with labels, Includes preprogrammed scenarios, Streaming voice headset, RF communications module, Bluetooth communications module, USB wireless router, Mother and baby battery chargers, User guide

Package inclusions:

•Palpation abdominal cover, Contraction abdominal cover, C-section abdominal cover, Postpartum abdominal cover, 2 umbilical cords, 2 precut umbilical cords,

	<p>Placenta, Epidural insert, C-section baby, 2 C-section abdominal skins, 2 C-Section uterine walls, Healthy PPH perineum, Midline episiotomy suture trainer, Mediolateral episiotomy suture trainer, 4th-degree episiotomy suture trainer, 2 suppositories, Artificial blood concentrate, Mineral oil lubricant, Hospital gown, baby head cap, receiving blanket and diaper, Manual blood pressure cuff, 2 baby delivery rings, Birthing baby cradle, Bladder, IV, and PPH filling kits, NIBP calibration kit, Antecubital vein replacement set, Spare cervix and birth canal</p> <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . Training 2-3 days.</p>
3.	<p>Maternal Fetal Birthing Simulator (HIGH FEDELITY MANIKIN) General Specifications: Fully Instrumented Mother and a Partially Instrumented Fetus, Muse Operating Software, MacBook Pro Instructor's Wireless Workstation, Touch-Pro Wireless Patient Monitor Computer, 10 Preconfigured SCEs, 4 SCE Development Licenses, First Year Full System Value Service Agreement Free, One Year of Training for end-users. Includes: Female Patient Module: turns your Birthing Simulator to a regular female patient simulator. Includes 5 Pre-Configured SCEs and a Non Gravid Abdomen Manikin/Simulator:</p> <ul style="list-style-type: none"> • Birthing Simulator no more than 176 cm long. • No more than 50 kg in weight • Fetus: Max 48 cm long and Max 2.5 kg in weight • Wireless patient monitor via All-in-One Touchscreen Computer • Must have Birthing Fetus, Placenta and Umbilical Cord • Separate Leopold's Maneuver Fetus • Prepartum Abdomen • Set of Static Cervices for vaginal exams • Postpartum Boggy/Contracted uterus with blood tank • Blood tank minimum of 1.8 liters to simulate Class III hemorrhage • Ability to Simulate Prepartum and Postpartum Scenarios • w/ Fully and partially inverted uterus, Uterine Inversion • Abdominal Cover for all fours birthing, Non-gravid abdomen • Upgradeable for Augmented Reality Training <p>Simulator Features: Breathing:</p> <ul style="list-style-type: none"> • Lung auscultation: anterior and posterior with individual lung control <p>Spontaneous breathing</p> <ul style="list-style-type: none"> • Spontaneous chest excursion synchronized with ventilation • Lung sounds can be independently controlled. Airway: Anatomically realistic upper airway. • Lungs allow for mechanical positive pressure ventilation support and allows ventilation modes (SIMV and CMV) • Airway Management and Ventilation, Bag-Valve-Mask • Articulated Mandible for jaw thrust maneuver • Detects right mainstem intubation • ET Tube, nasal-pharyngeal and oropharyngeal airways • Must feature an articulating mandible <p>Blinking eyes that react to light Light reactive and blinking can be controlled via software options</p> <ul style="list-style-type: none"> • Seizures via rhythmic movement of arms and rapid blinking • Eyes must be able to reflect the following conditions: Jaundice, Hemorrhage, Keyhole Pupil Cataract, Bloodshot, Droopy Eyelid <p>Urological:</p> <ul style="list-style-type: none"> • Urine output <p>Sounds:</p> <ul style="list-style-type: none"> • Heart, bowel and breath sounds, Vocal Sounds

<ul style="list-style-type: none"> • Other sounds can be provided via wireless microphone • Two-way communication • Audible breathing sounds • Vocal sounds can be synchronized with uterine activity <p>Vascular Access:</p> <ul style="list-style-type: none"> • IV insertion supported in right arm: cephalic, basilic and antebrachial veins • IO site on anterior tibia • Right jugular IV line for infusions <p>Articulation:</p> <ul style="list-style-type: none"> • Range of motion in wrists, elbows, hips, knees and ankles. <p>Obstetrical:</p> <ul style="list-style-type: none"> • Integrated modeling of maternal-fetal physiology measurements • Maternal aesthetics build from real patient <p>Realistic birth canal and vulva/perinium supports accurate fetal descent and rotation</p> <ul style="list-style-type: none"> • Birthing Positions: lithotomy, sitting and all fours • Rectal suppository administration <p>Prepartum:</p> <ul style="list-style-type: none"> • Vaginal examinations of cervix, fetal station and position • Static cervixes representing various stages of dilation, from closed to 5cm, effacement from 0 to 90% • Leopold's maneuver can be performed to determine presentation and position of fetus • Epidural port <p>Intrapartum:</p> <ul style="list-style-type: none"> • Realistic and palpable uterine contractions detected by palpating the fundus, Rate and duration of contractions can be controlled via software, Trendelenburg position with detection, Left lateral tilt with detection, Vertex and Breech delivery, Application of McRoberts Maneuver with observable pelvic tilt, supported and logged, Suprapubic pressure support and detection with palpable symphysis pubis, Supports delivery of posterior arm for shoulder dystocia • Zavanelli maneuver with detection • C-section team training support • Rotation of anterior and posterior shoulder detected and resolves shoulder dystocia when performed correctly via Rubin II and Woods' Screw Maneuvers • Forceps application • Vacuum extraction without need for fetal cap • Fetal hear sounds: 5 locations depending upon fetal presentation • Placenta: intact/fragmented with realistic color/texture/21flexibility Placenta delivery <p>Fetus:</p> <ul style="list-style-type: none"> • Birthing fetus without unrealistic connection ports in the head and/or buttocks • Articulated fetal body neck, shoulders, elbows, hips and knees • Fetus has lateral neck movement • Clinically accurate fetal size at 5th percentile of WHO growth chart All scenarios provided must be evidence based and have the following: <ul style="list-style-type: none"> o Synopsis <ul style="list-style-type: none"> • Scenario Programming of Patient History • SBAR Handoff Report • Healthcare Provider Orders Learning Objectives • Learning Performance Measures • Preparatory Questions • Suggested Equipment and Supplies needed of <p>Facilitator Notes</p> <ul style="list-style-type: none"> o Debriefing Points
--

	<ul style="list-style-type: none"> o Teaching Q & A o Clinical References (Evidence Based) <p>Included Scenarios: 15 Scenarios</p> <p>Birth Scenarios</p> <ol style="list-style-type: none"> 1. Normal delivery 2. Instrumented vaginal delivery 3. Fetal Tachycardia due to maternal pyrexia 4. Breech Delivery 5. Fetal Central Nervous System Depression by narcotic given to mother 6. Shoulder Dystocia 7. Major Post-partum Hemorrhage due to uterine atony 8. Maternal Cardio-respiratory Arrest 9. Eclampsia 10. Umbilical Cord Prolapse <p>Non-Birthing Scenarios</p> <ol style="list-style-type: none"> 7. Chronic Heart Failure Exacerbation 12 Acute Respiratory Distress Syndrome 13.Sepsis with Hypotension 14. Brain Attack w/ Thrombolytic Therapy 15.Motor Vehicle Collision with Hypovolemic Shock <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . Training 2-3 days.</p>
4.	<p>Adult Training Manikin/ Geriatric Manikin</p> <p>Lightweight – approximately 28 lbs.</p> <ul style="list-style-type: none"> •Full-size adult manikin – measures 58 in. •Overall female appearance •Converts to male with removal of wig and attachment of male genitals <p>Visual Inspections:</p> <ul style="list-style-type: none"> •Normal and Cancerous Mole •Stage 1 Sacral Ulcer •Dilated and Constricted Pupils •Reddened Skin Folds <p>Patient Care Simulation:</p> <ul style="list-style-type: none"> •Bandaging and Wound Dressing •Bed Baths <p>Clothing Changes Patient Care Simulation:</p> <ul style="list-style-type: none"> •Bandaging and Wound Dressing •Bed Baths •Clothing Changes •Denture Placement and Removal, Upper and Lower •Ear Canal Irrigation, Optic Drops, and Hearing Aid Placement •Eye Irrigation •Finger and Toe Manipulation with Bandaging – Flexible and Individually Molded •Hair Care Washing and Combing •Intramuscular Injection Sites: Arm, Thigh, Buttock •Oral and Nasal Hygiene: Lavage, Gavage, and Suctioning •Ostomy Care: Ileostomy and Colostomy Tissue •Maintenance and Appliance Application, Lavage, and Suctioning •Patient Positioning •Patient Transfer Techniques •Pericare •Tracheostomy Care, Lavage and Suctioning •Gastrostomy Procedures, Lavage and Gavage •Ankle – Eversion, Inversion, Dorsiflexion, Plantarflexion •Elbow – Extension, Flexion, Pronation, Supination •Fingers – Abduction, Adduction, Flexion (soft, lifelike material)

	<ul style="list-style-type: none"> •Hip – Abduction, Adduction, Rotation, Hyperextension •Knee – Extension, Flexion •Enema Administration, Female Only •Urinary Catheterization, Female and Uncircumcised •Neck – Rotation, Hyperextension, Lateral Flexion •Shoulder – Abduction, Adduction, Rotation, Hyperextension •Toes – Abduction, Adduction, Flexion (soft, lifelike material) •Trunk – Rotation, Hyperextension •Wrist – Flexion, Hyperextension, Radial Flexion, Ulnar Flexion <p>Realistic Patient Positioning, Flexibility, & Superior Range of Motion Male</p> <ul style="list-style-type: none"> • Pap Smears and Douching • Prostate Exam, Stage B • Nasogastric Tube Placement • Internal Fluid Reservoirs <p>Warranty: 2 years on parts and service. Preventive Maintenance: once a year for 5 years</p> <p>Training 2-3 days.</p>
5.	<p>Human Whole Body Plastinated Specimen</p> <p>Dry & odorless, No oil, water, molds or any leaks present in the specimen to show signs of decomposition</p> <p>Left Half showing superficial muscles and other right half in deep muscle dissection. Front and back. Synopsis: Skeletal muscle is an important portion of locomotive system and is also the emphases of teaching. After delicately dissected, the specimen displays the skeletal muscle of the whole human body and some cutaneous nerves and superficial arteries which are remained simultaneously. One side of the Specimen displays superficial layers muscles and the other side displays Deep layer muscles. By means of the contrast between the two sides, students can deeply understand and grasp the position and relationship of the muscles.</p> <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years .</p> <p>“Terms of reference for plastinated specimens”</p> <ol style="list-style-type: none"> 1. Plastinated specimens are preserved using specialized silicone or compound for whole body or regional parts. 2. Plastinates should be cleanly dissected to reveal specific anatomical parts and not twisted, deformed or distorted. 3. Plastinates should have a life-like natural color. 4. Plastinates should be completely dry to touch not gluey, sticky or watery. 5. Plastinates should be odorless and should not emit any odor/ foul smell or when stored in glass containers 6. Plastinates should be non toxic and does not possess any allergic reactions to the eyes, nose, skin etc. 7. Plastinates should be portable i.e not heavy 8. Distributor should replace the plastinated specimen when it does not conform to the conditions above. 9. Plastinates should have a 3-dimensional feature 10. Plastinates can be stored in free air (not airconditioned) indefinitely. Humidity ranges below 60% and can withstand temp range of 20-30 degc 11. Distributor should render free aftersales and preventive maintenance service for at least 5 years. 12. Distributor should be able to provide repair when specimens are damaged. 13. The distributor must show an actual demo unit during post qualification
6.	<p>Advance Patient Training Arm</p> <p>Subcutaneous injection sites on the volar forearm and lateral upper arm</p> <p>Intramuscular injection site on the upper arm</p>

	<p>Arterial system including the radial and brachial arteries Suture and incision sites on both the upper arm and forearm Hemodialysis site on the forearm Arterial and Venous insert for IV training, venipuncture practice, blood draw exercises, AV anastomosis, and placement of AV grafts. This multi-layer surgical insert includes the skin, subcutaneous tissue, muscle, radial artery, and radial vein AV fistula insert that simulates a healed fistula for hemodialysis exercises. Multi-layer bicep insert that includes the skin, subcutaneous tissue, and muscle and allows incision practice and suturing exercises. Durable skin that can be pierced in excess of 200 times with a 20 or 22 gauge needle. Realistic tactile feedback for both surgical and arterial & venous stick exercises. Adjustable heart rate and pulse strength simulating a heart rate from 10 BPM to 150 BPM. Cephalic (antecubital), Basilic, Radial, and Ulnar veins, as well as the radial and brachial arteries for infusion and blood draw training. Rotating arm allowing dorsal and volar access along the length of the IV training arm. Varying vessel palpability to simulate collapsed or bulging vessels. Ease of assembly. Latex-Free New proprietary materials replicate the skin, subcutaneous, and muscle layers at all surgical sites. Upgraded inserts, skins, and vessels for improved tactile feedback Micropump embedded within the shoulder generates variable heart rates and pulse strength. Quiet and compact, thereby improving the portability of the unit. Easy to replace plug-and-play inserts to allow quick change-out between procedures. Latex-free vessels with improved access for hassle-free replacement</p>
7.	<p>CPR Manikins (set of 3 manikins-Adult, Child & Infant) Features: • CPR feedback monitor that meets 2015 AHA CPR guidelines for Adult Manikin. • Available in Adult, Child, and Infant models. • Realistic looking and performing manikin is more acceptable to both instructors and students. • Far easier and faster to carry and set up at training sites. • Patented face/head tilt simulates how an actual victim’s head would move if he required CPR and teaches students how to properly open the airway. • Easily inserted face-shield/lung-bag provides visual chest rise to ensure students receive the most accurate and realistic training. The face shield is secured so it won’t slip and there’s no need for adhesives. Prestan Manikins are not made with natural rubber latex 2 Extra Compression Springs (Hard and Soft) Training Pads, Airway Lubricant, Eye Set, Kneel Mat, Jacket, Pants Manikin Wipes, USB Cable Plus a ShockLink with Defibrillator Training Pads for ShockLink ShockLink Remote Control ShockLink Pouch and User Guide Includes 1 x Adult Manikin with CPR monitor 1 x Child Manikin with CPR monitor 1 x Infant Manikin with CPR monitor 10 x Adult Face Shield Lung Bags 10 x Child Face Shield Lung Bags 10 x Infant Face Shield Lung Bags 1 x Nylon Carrying case Warranty: 1 year</p>
LOT 2 – HOSPITAL EQUIPMENT	
8.	<p>BABY BASSINET with stand single Bassinet stand with acrylic glass basket Bassinet stand dimension: 83 x 46 x 64.5cm Frame leg and support stainless steel</p>

	<p>Shelf-stainless steel Mattress of the bassinet is ordinary foam with cloth cover With four lockable caster wheels at 4 inches in size Basket dimension: (H) 22cm x (L) 75.5cm x (W) 40cm Warranty: 1 year</p>
9.	<p>EMERGENCY BED STRECHER Operated by Handle. When you turn it right, it is up, turn left, it is down. The Central-Lock System and Swivel Castor can help the User easy to control and stop. Wheel Size: Φ150mm Double Sided. High position: 192x61x83cm Low position: 192x63x54cm Size of the castors: Dia.150mm Max angle of the back: 65 0 Self-weight: 85kgs Load bearing: 159kgs Packing Quantity: 1 Size(LxWxH): 198x73x64cm Warranty: 1 year</p>
10.	<p>Medical Emergency Crash Cart 1. Mainly of aluminum, steel and ABS engineering plastic; bear Four columns to bear weight; 2. Upper part: ABS integrated board molding by injection one-time, with handralls on both sides, small items won' t slip; With transparent soft glass on the table effectively prevents scratches 3. Front: Push and pull smoothly, the Centralized Lock, Key can be folded to ensure the safety of the items inside the drawer; The drawer handle is dovetail shaped, which can effectively prevent dust from entering. Five drawers, which are divided into two small, two medium, and one large drawers. Each drawer inside with 3*3 dividers, can be separated freely. Label card holder is slot type design, to prevent liquid and dust into holder. The label card holder is inclined upward for convenient looking, which is according to the ergonomic principle. 4. Left side: Difillatoar shelf, which is used to store the defibrillator. It has fixed magic straps and does not fall off. Pull-out sliding writing frame improves space application efficiency Movable File Container enables easier access to patient information 5. Right side: Equipped with lifting infusion stand, convenient and easy to operate; One 2L needle disposal holder in the utility basket, Double ABS dust basket for waste sorting will effectively eliminate the smell of garbage; 6. Back: -CPR Board, for acute stroke patients - Oxygen Tank Holder with belt attached to the body of the trolley. Does not drop when moving. -power outlet & hook with 5m wire. 7. Bottom: -luxury universal insert noiseless casters, two with brake, casters are made of high-strength polyurethane. -Anti- static, anti-hair winding, portable and flexible; 750*475*930mm 5 Layers Packing Size: 760*540*920 mm Rack of multi bin container Disposable lock with apple-shape, IV Pole, Defibrillator shelf Warranty: 1 year</p>
11.	<p>INFUSION PUMP (1) and SYRINGE PUMP(1) Extendable from 3 up to 6 or 9 Pump Slots • by simply connecting 2 or 3 Standard Racks (TE*RS700N) or by adding 1 or 2 Rack Extension Devices to the Communication Rack Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of</p>

	<p>accidental bolus infusion Reliable high accuracy combination of pump & sets* Stable infusion thanks to dedicated set Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Luer-Lock syringes Infusion Sets Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of Terumo pump & sets* Stable infusion thanks to dedicated set * Flow rate accuracy: within $\pm 5\%$ Available in different sizes Transparent barrel Bold and easy legible scale markings No Latex & PVC components Silicon coating Double contact gasket Infusion Sets Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of pump & sets* Stable infusion t* Flow rate accuracy: within $\pm 5\%$ Available in different sizes Transparent barrel Bold and easy legible scale markings No Latex & PVC components Silicon coating Double contact gasket Warranty: 1 year</p>
12.	<p>INFANT DIGITAL WEIGHING SCALE Special feature Backlit Display Material: Plastic Item weight: 8 Pounds Small Footprint. Accuracy to 1 2 ounce. Detachable tray for easy cleaning and portability. Ergonomically designed for both ease of use and infant safety. features: metric lockout, integral measuring tape (up to 22 inch/56 cm), 30 lb x 0.1 oz / 15 kg x 0.005 kg capacity, locking pin for easily removing the tray for cleaning, 1 inch (25 mm) high LCD, and RS232 serial output for connectivity with a printer or PC. The spacious weighing tray measures 26 in W x 12 in D / 66 cm W x 31 cm D. The scale's ZERO button removes blanket tare weights. Your choice of rechargeable NiCad batteries (not included), standard Alkaline batteries (not included), or UL-listed AC adapter (included). The 6745 pairs perfectly with the Detecto stainless steel baby scale cart. Warranty: 1 year</p>
13.	<p>12 LEAD ECG MACHINE Channels Record & print 12 channels of data at a time, fastest speed Standard 12 leads; Cabrera; Nehb, VCG 9 Inches, 15° angled display for clear visualization Touchable, time-saving operation quick input of patient data Rechargeable, 3-hour continuous work Local Storage: 3,500 files Thermal Paper:Z-fold paper: 210*140mm Economic Recording Mode Weight: 3.4KG Accessories in the Package ECG Acquisition Box with Lead Wire Z-fold Thermal Paper (160 Pieces: 112mm × 140mm) Power Cord ×2 (US & EU Standard) & AC Power Adapter Equipotential Grounding Cable Warranty: 1 year</p>

14.	<p>Orthopedic bed with balkan frame, 3 cranks</p> <ul style="list-style-type: none"> · Firm mattress · Fracture board · Bed elevator/ shock block (optional) · Balkan frame (cross bar, curve bar, vertical, horizontal, diagonal bars) · 3 pulleys · 3 clamps · Overhead trapeze <p>Dimensions: L2240xW1060xH470=680mm Function : 3 functions Material : Cold rolled steel, HDPE, Aluminum alloy Side rails : ABS Side Rails Material: Cold rolled steel, HDPE, Aluminum alloy Dimensions : L2240xW1060xH470=680mm Bed platform : 1980x900mm Backrest : 0-75deg +-5deg Legrest : 0-45deg +-5deg Height adjustment : 470-680+-10mm Brake system : Central brake Sticker color : Wood Grain/Lake Blue Side rails : ABS side rails Orthopedics traction shelf : Aluminum Alloy 2023”793”1503mm</p> <p>Features</p> <ol style="list-style-type: none"> 1. Blue liquid angle indicator on backrest and leggiest side rails 2. Clear acrylic head card for patient name display, four infusion holes, four drainage hooks around the bed 3. Mechanical crank operation with overload protection 4. The screw handle is foldable under the bed footboard 5. Four 125mm diameter castors with central brake 6. Orthopedics traction is made of molding aluminum alloy 2023*793*1503mm <p>Warranty: 1 year</p>
15.	<p>Balance Skeletal Traction Equipment</p> <ul style="list-style-type: none"> · Thomas splint (with half ring) · Pearson’s attachment · Rest splint · Steinmann’s pin holder · Braun-Bohler splint · Slings with clips or pins · 3 sash cords (thigh rope, traction rope, suspension rope) · 2 weight bags <p>Traction weight (10% of patient’s body weight) Suspension weight (1/2 of traction weight) · Foot pedal or foot board</p> <p>Warranty: 1 year</p>
16.	<p>Limb (Foot) use for Balance Skeletal Traction (1 adult and 1 Pedia Limb Warranty: 1 year</p>
17.	<p>OR LIGHT SINGLE HEAD</p> <p>max illumination power 160,000lux diameter of lighthouse 670mm color temperature 3000k-5000k color rendering index 96% dimming range 1-100% light field diameter d'10 24-36cm/9.5"-14.2" light field diameter d50 12-28cm/5.1"-11" number of leds 60leds lifetime of led >60,000hours degree of turbulence (standard value 37.5% max) 29% depth of volume of light L1 + L2 at 20% 130cm/51.2"</p>

	<p>depth of volume of light L1 + L2 at 60% 76cm/30" irradiance at nominal illuminance 510w/m2 handle holder made of ABS plastic flexible arms that provide comfortable positioning there is a laser marking feature to mark the midpoint of the central illumination intensity endo light mode/standby mode can be switch by a single button Warranty: 1 year</p>
18.	<p>OPERATING ROOM TABLE Electro Hydraulic OR Table Easy to handle modular design for different surgical disciplines Long and strong rail line for accessories T base design for maximum approach Hand control and column control units Battery back up: up to 30 full movements</p> <p>Technical Features: Sliding Movement: 300mm Lateral Tilt Movement: +20 degree Celsius / -20 degree Celsius Back Rest Movement: +80 degree Celsius / -45 degree Celsius Trendelenburg: +30 degree Celsius / -80 degree Celsius Reverse Trendelenburg: -30 degree Celsius / -90 degree Celsius Leg Rest Up/Down: +30 degree Celsius / -80 degree Celsius Head Rest Movement: +30 degree Celsius / -90 degree Celsius Flex/Reflex Position: +225 degree Celsius / -100 degree Celsius Width x Length: 560mm x 2200mm Height (Min): 595mm Height (Max): 1035mm Mattress Thickness: 70mm Castor Diameter: 100mm Maximum Patient Weight: 360kg Stainless Steel T Base Geopel Leg Holder -made of high-quality stainless steel -pair, including clamp -height adjustable -rotatable and tillable via ball-join Warranty: 1 year</p>
19.	<p>Digital Binocular Biological Microscope Optical System: Infinite Optical System Eyepiece: WF10x / 20mm, ANTI FUNGUS with pointer Viewing Head: Seidentopf Binocular Head, inclined at 30°, Interpupillary 47-78mm ANTI FUNGUS Objective: Infinite Semi-Plan Achromatic Objectives 4x, 10x, 40x, 100x ANTI FUNGUS Nosepiece: Backward Quadruple Nosepiece Stage: RACKLESS Double Layers Mechanical Stage 150mm×139mm, Moving Range 75mm×52mm Condenser: Abbe Condenser NA1.25 Focusing: Coaxial Coarse and Fine Adjustment, Fine Division 20mm, Moving Range 0.002mm Illumination: S-LED Illumination with Rechargeable Battery, Brightness Adjustable, can be connected to an LCD monitor or HD TV via HDMI cable and operated independently without connecting to PC. The image/video capture and operate can be controlled by mouse, so no shaking when you take images and videos. It also can be connected to a PC via USB2.0 cable and operate with the software. Features</p>

	<p>Use mouse to control the Camera. Record image and video to SD card. High frame rate of 15fps. The functions inside camera (Cloud 1.0)The less Icons the Better. Set Exposure Time Capability. 3D Noise Reduction. 1080P Video Recording. Get more Details with ROI Magnification Function. Image Comparison Function. Browse Captured Images PC software Warranty: 1 year</p>
20.	<p>Electric Wheelchair Total length of wheelchair: 920MM Total height of wheelchair body: 850MM Seat width: 460MM Rear wheel size: 12 inch Front wheel size: 8 inch Frame material: steel Motor power: 250W (2) 60HZ Climbing capacity: 35 ° Net weight 30KG Both sides open and close armrest Breathable cushion High-strength ABS anti-lashing rear wheel Thickened footrest Thickened steel Adjustable seat belt Back-up reminder Large capacity back pocket Anti-dumping rear wheel EPBS braking system without skidding Warranty: 1 year</p>
21.	<p>WEIGHING SCALE WTH HEIGHT MEASUREMENT Heavy-duty solid stable platform Dual-reading die-cast weight beam which may be read from either side of the scale Durable steel construction with an electrostatic powder paint finish for optimum quality. Capacity may be increased by using an optional counterweight Capacity: 400 lb x 4 oz , 180 kg x 100 g, 400 lb x 4 oz / 175 kg x 100 g Scale Height: 59 in / 150 cm Height Rod (on select models - 2392 excluded): 30 to 78 in / 76 to 200 cm - 1/4 in / 1/2 cm increments Platform Size: 10.5 in W x 14.5 in D / 27 cm W x 37 cm Warranty: 1 year</p>
22.	<p>HOSPITAL BED Electric multi-functional nursing bed medical bed hospital bed turn over bed rehabilitation lift bed Intelligent remote control operation (also supports manual) Sending shampoo basin, universal wheel, urinal, infusion rack, dining board International certified motor (38DB) low noise Can be timed to turn over (30-45 minutes once) Electric start back One piece reset Electric lift leg (0-25°) Electric drop leg (0-85°) Electric turning (0-45°)</p>

	<p>Electric helper for defecation Head washing function Foot washing function Infusion function Meal function Mobility function Automatic lifting of the guardrail Anti-wetting mattress Size:2060*1030*480mm Bed size:1950*900mm. Warranty: 1 year</p>
23.	<p>36" Discover Headwall Package PRODUCT FEATURES •The Functioning Wall Mounted 36" Discover Headwall Package is the most popular training headwall. It simulates oxygen, air and suction with one source of air by connecting to an external Silent Air Compressor. The Discover headwall package is designed for simulation labs so it is easy to install and maintain, and is available in a variety of sizes and color options, Discover size: 39.38" H x 19.5" W x 4.75" D, Wall mounted: Mounting strip included, Nurse call facade with code blue button and hand pendant TECHNICAL DATA • Size: 36", Wall Mounted – Includes mounting strip, Ohmeda gas outlets, Oxygen, Air, Vacuum, Duplex receptacle, Vacuum slide, Nurse call 3 gang cover, Includes Complete Accessories Package •Oxygen Flow Meter, Medical Air Flow Meter, Analog Vacuum Regulator, Suction Canister, Humidifiers, Nasal Cannulas, Suction tubing, DISS Hose and Fittings Warranty: 1 year</p>

Note: Bidders must state in the Statement of Compliance either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidders statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the provision of **ITB** Clause 4.

Section VIII. Bid Documents Checklist

This Bid Documents Checklist is provided to guide the Bidder in preparing his/her bid. The checklist may be used by the Bidder to verify if the Bid includes all the prescribed documents.

The Bidder, in submitting the required documents, must use the prescribed forms found in Section X. Bidding Forms. However, should a bidder choose to use a different formatting style for a required document, the bidder must ensure that the substance in the form given in Section X for that particular document is substantially captured in the equivalent document.

I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

Legal Documents

- (a) Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR;

Technical Documents

- (b) Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid; and
- (c) Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents; and
- (d) Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission or Original copy of Notarized Bid Securing Declaration; and
- (e) Conformity with the Technical Specifications, which may include production/delivery schedule, manpower requirements, and/or after-sales/parts, if applicable; and
- (f) Original duly signed Omnibus Sworn Statement (OSS) and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.

Financial Documents

- (g) The prospective bidder's computation of Net Financial Contracting Capacity (NFCC) or A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation.

Class "B" Documents

- (h) If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.

II. FINANCIAL COMPONENT ENVELOPE

- (i) Original of duly signed and accomplished Financial Bid Form; and
- (j) Original of duly signed and accomplished Price Schedule(s).

Section IX. Bidding Forms

TABLE OF CONTENTS

A. Eligibility and Technical Documents	
b. Statement of all on-going government and private contracts, including contracts awarded but not yet started, if any	33
c. Statement of Single Largest Completed Contract (SLCC), similar to the contract to be bid, in accordance with ITB Clause 5.3	34
d. Bid Securing Declaration or Bid Security, in the form, amount and validity period, as prescribed in ITB Clause 14.1	35
e. Bidder's Compliance to the Technical Specifications	37
f. Omnibus Sworn Statement, which shall be duly notarized	53
g. NFCC Computation	55
B. Financial Documents	
i. Financial Bid Form	56
j. Schedule of Prices	57

[Bidder's Letterhead]

[Date]

To: Tarlac State University
Re: Invitation to Bid No.

Statement of Single Largest Completed Contract Similar to the Contract to be Bid

Row 1: Name of Contract Row 2: Location	Contract Price	Row 1: Procuring Entity Row 2: Address Row 3: Contact Person/Tel. No.	Description of Goods	Date of Award	Date Completed

Attached herewith are the following documents: Contract Agreement, Notice of Award, Notice to Proceed, Official Receipt/Invoice, Certificate of Final Inspection, and Certificate of Acceptance, as evidences in support of the foregoing information.

I/We certify that the foregoing information and all of the supporting documents are true and correct.

[Signature]
[Name of Bidder or Authorized Representative]
[Position or Title]

Bid-Securing Declaration

Republic of the Philippines
City/Municipality Of _____) S.S.

x-----x

Invitation to Bid [*Insert reference number*]

To: **Tarlac State University**

I/We, the undersigned, declare that:

1. I/We understand that, according to your conditions, bids must be supported by a Bid Security, which may be in the form of a Bid-Securing Declaration.
2. I/We accept that: (a) I/we will be automatically disqualified from bidding for any contract with any procuring entity for a period of two (2) years upon receipt of your Blacklisting Order; and, (b) I/we will pay the applicable fine provided under Section 6 of the Guidelines on the Use of Bid Securing Declaration, within fifteen (15) days from receipt of written demand by the procuring entity for the commission of acts resulting to the enforcement of the bid securing declaration under Sections 23.1(b), 34.2, 40.1 and 69.1, except 69.1 (f), of the IRR of RA 9184; without prejudice to other legal action the government may undertake.
3. I/We understand that this Bid-Securing Declaration shall cease to be valid on the following circumstances:
 - a. Upon expiration of the bid validity period, or any extension thereof pursuant to your request;
 - b. I am/we are declared ineligible or post-disqualified upon receipt of your notice to such effect, and
 - i) I/we failed to timely file a request for reconsideration or
 - ii) I/we filed a waiver to avail of said right;
 - c. I am/we are declared as the bidder with the Lowest Calculated Responsive Bid, and I/we have furnished the performance security and signed the Contract.

IN WITNESS WHEREOF, I/We have hereunto set my/our hand/s this ____ day of [*month*] [*year*] at [*place of execution*].

[*Signature*]
[*Name of Bidder's Authorized Representative*]
[*Signatory's legal capacity*]
Affiant

SUBSCRIBED AND SWORN to before me this ___ day of *[month]* *[year]* at *[place of execution]*, Philippines. Affiant/s is/are personally known to me and was/were identified by me through competent evidence of identity as defined in the 2004 Rules on Notarial Practice (A.M. No. 02-8-13-SC). Affiant/s exhibited to me his/her *[insert type of government identification card used]*, with his/her photograph and signature appearing thereon, with no. _____.

Witness my hand and seal this ___ day of *[month]* *[year]*.

[Name and Signature of Notary Public]

Serial No. of Commission _____

Notary Public for _____ **until** _____

Roll of Attorneys No. _____

PTR No. __, *[date issued]*, *[place issued]*

IBP No. __, *[date issued]*, *[place issued]*

Doc. No. _____

Page No. _____

Book No. _____

Series of _____.

[Bidder's Letterhead]

[Date]

To: Tarlac State University
 Re: Invitation to Bid No.

Compliance to the Technical Specifications

Item No.	Description	Bidder's Compliance <i>State "Comply" below if your offer is complying with the specifications, otherwise state "We are offering the same goods with the following specifications [State the complete specifications of the alternative offer]</i>
1.	<p>PEDIATRIC SIMULATOR (HIGH FEDELITY)</p> <p>General</p> <ul style="list-style-type: none"> • Height: 44 inches • Tetherless and wireless; fully responsive during transport • The internal rechargeable battery provides hours of tetherless operation • Smooth and supple full-body skin with seamless trunk and limb joints • Realistic joint articulation: neck, shoulder, elbow, hip, and knee • Palpable bony landmarks • Forearm pronation and supination • Supports common patient positions including Fowler's, supine, and sitting • Male/female patient conversion • Tablet PC preloaded with UNI® included • Includes 10 preprogrammed SLEs and facilitator's guidebook <p>Neurological</p> <ul style="list-style-type: none"> • Active robotics simulate lifelike facial expressions including: anger, transient pain, ongoing pain, amazement, Quizzical, crying, yawning <p>Preprogrammed emotional states automatically express associated verbal and non-verbal cues without manual input: Worried, Anxious, Lethargic, Distracted</p> <ul style="list-style-type: none"> • Abnormal eye and eyelid movements: crosseyed, nystagmus, eyelid twitching, eyelid droop, Programmable crying/ tears release real fluid, Wireless streaming voice: listen to participants respond in real-time, Real-time voice modulation effects, Automatic jaw movement synchronized with speech, Automatic jaw movement synchronized with speech, Seizures with selectable intensity levels, with prerecorded speech responses • Real CO2 exhalation: supports etCO2 monitoring using real sensors and monitoring devices, Selectable normal and abnormal sounds: upper right, front and back; upper left, front and back; lower right, back; and lower left, back, uses "Real" mechanical ventilation support, Real-time ventilation feedback, Visible chests rise during BVM ventilation, Chest tube insertion: left midaxillary hemothorax site features palpable bony landmarks, realistic skin for cutting and suturing, tactile pleural pop, and fluid drain, Needle 	

	<p>decompression site features realistic tactile feedback and audible hiss, Needle decompression and chest tube insertion</p> <p>Cardiac</p> <ul style="list-style-type: none"> •Includes comprehensive library of ECG rhythms with customizable beat variations, Independent normal/abnormal heart sounds at aortic, pulmonic, and mitral sites, Supports ECG monitoring using real devices •Supports ECG-derived respiration monitoring (EDR), Ecpr™ Real-time quality feedback and reporting •Defibrillate, cardiovert and pace using “real devices” and energy •Circulatory, Visible cyanosis, redness, pallor, and jaundice, Supports capillary refill time testing above the right knee; test detection and logging, Palpable pulses: bilateral carotid, brachial, radial, femoral, and pedal, Blood pressure-dependent pulses, Supports blood pressure monitoring using a “real NIBP” cuff and monitor, SpO2 monitoring using real devices •Vascular Access, Bilateral forearm IV access supports sampling and continuous infusion, Intraosseous infusion site at right proximal tibia, Real glucose test readings via finger-stick •Gastrointestinal, Patent esophagus, Gastric distension during excessive PPV, Bowel sounds in four quadrants, Interchangeable male/female genitalia, Supports urinary catheterization with fluid return, Programmable urinary output <p>Airway</p> <ul style="list-style-type: none"> •Anatomically accurate oral cavity and airway, Supports nasotracheal/ orotracheal intubation with standard instruments including endotracheal tubes and supraglottic airway devices, Tracheal intubation detection, Head tilt, chin lift, jaw thrust, Supports esophageal intubation, NG/OG tube placement, Supports bag-valve-mask ventilation, Realistic surgical trachea permits tracheostomy, cricothyrotomy, and retrograde intubation, Programmable difficult airway: laryngospasm and tongue edema, Selectable normal and abnormal upper airway sounds <p>Breathing</p> <ul style="list-style-type: none"> •Spontaneous breathing and selectable normal and abnormal respiratory patterns •Variable respiratory rates and inspiratory/expiratory ratios •Programmable unilateral chest rise and fall •Unilateral chest rise with right mainstem intubation <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . Training 2-3 days.</p>	
2.	<p>Maternal and Neonatal Birthing Simulator (ULTRA HIGH FIDELITY MANIKIN)</p> <p>GENERAL SPECIFICATION</p> <ul style="list-style-type: none"> •Tetherless and wireless; fully responsive during transport, Wireless control at distances up to 300 ft. Internal rechargeable battery provides up to 10 hrs. of tetherless operation, Smooth and supple full-body skin with seamless articulating joints, Lifelike joint articulation: neck, shoulder, elbow, wrist, hip, knee, and ankle, Internal high-capacity fluid reservoirs, Maternal® Fetus-Newborn wireless link capability, Includes touchscreen tablet PC preloaded with software, Includes touchscreen virtual patient monitor and CTG <p>TECHNICAL DATA SHEET</p> <p>Processor : 12th Gen Intel® Core™ i5-12500 (18 MB cache, 6 cores, 12 threads, 3.00 GHz to 4.60 GHz Turbo, 65 W)</p>	

Operating System : Windows 11 Pro
 Graphics Card : Intel® Graphics
 Monitor : OptiPlex All-in-One Non-Touch Panel
 Memory * : 8 GB, 1 x 8 GB, DDR4
 Hard Drive : 256 GB, M.2 2230, PCIe NVMe, SSD, Class 35
 Keyboard : Wireless Keyboard and Mouse
 Ports on the Side : 1 USB 3.2 Gen 2x1 Type-C® port
 1 USB 3.2 Gen 1 port with PowerShare
 1 Universal audio port
 Rear : 2 USB 3.2 Gen 2 ports
 2 USB 3.2 Gen 1 ports with Smart Power On
 1 RJ45 Ethernet port
 1 Line-out audio port
 1 DisplayPort++ 1.4a/HDCP 2.3 port
 1 HDMI-IN—HDMI 1.4a port
 (with UMA configuration-65 W)
 1 HDMI-OUT—HDMI 2.0 port
 (with UMA configuration-65 W)
 Slots : 1 SD-card slot (side)
 1 M.2 2230 slot for Wi-Fi and Bluetooth card
 1 M.2 2230/2280 slot for SSD (35 W)
 2 M.2 2230/2280 slot for SSD (65 W)

NEUROLOGICAL

•Programmable blinking rate, pupil response, and bilateral and unilateral eye movement, Interactive eyes can follow a moving object, Seizures with selectable intensity levels, Wireless streaming voice: be the voice of Victoria and listen to participants’ responses via a wireless headset, Record and playback vocal responses in any language

AIRWAY: Head tilt/chin lift/jaw thrust, Oral or nasal intubation, Difficult airway: laryngospasm, tongue edema, and pharyngeal swelling.

BREATHING: Spontaneous breathing; selectable respiratory patterns and lung sounds, Programmable respiratory rates and inspiratory/expiratory ratios
 Realistic chest rise during assisted ventilation, Ventilation performance, real-time monitoring, reporting, and logging,

CARDIAC: Real-time compression and ventilation performance, Effective compressions generate palpable pulses and ECG artifacts. Real-time 4-Lead electrocardiogram monitoring using real ECG devices, Defibrillation, cardioversion, and pacing using real energy, Realistic heart sounds, Programmable heart rate synchronized with ECG and pulses, Virtual 12-lead dynamic ECG rhythms

CIRCULATION

•Bilateral palpable carotid, radial, and brachial pulses are synchronized with heart rate and blood pressure, Monitor oxygen saturation using real oximeters, Measure blood pressure using a real manual or automatic blood pressure cuff, Bilateral venous access, Automatic drug recognition detects medication type, dose, and rate injected into the lower right arm, SubQ and intramuscular injection sites for placement exercises

OBSTETRICS

•Maternal-fetal physiologic link, Supports Leopold’s Maneuvers and external cephalic version, Supports fetal monitoring using real devices, Precision Delivery System: computer-controlled fetal

	<p>descent and cardinal movements, Programmable normal, breech, shoulder dystocia, instrument assisted delivery, and C-section, Software activated self-lubricating birth canal, Realistic placenta with detachable fragments and umbilical cord</p> <p>OBSTETRICS</p> <ul style="list-style-type: none"> •Epidural placement and needle detection; palpable anatomical landmarks and skin layers, Intrapartum bleeding, Force sensor reports traction and torsion applied to the fetus in real-time, McRoberts maneuver, suprapubic pressure, Zavanelli, Woods' Screw, and "hands and knees" position detection, Births full-term neonate of realistic size and weight, Active neonate: programmable heart and respiratory sounds, crying, cyanosis, and movement, Replaceable episiotomy inserts support suturing, Multi-layer fourth degree episiotomy with hemorrhaging vaginal sidewall tears and cervical lacerations, Programmable uterine firmness and hemorrhaging supports sutures and balloon tamponade, Programmable hemorrhage flow control <p>GASTROINTESTINAL</p> <ul style="list-style-type: none"> •Selectable bowel sounds •Internal fluid bladder with urethra for Foley catheterization exercises •Rectum with suppository placement sensor <p>Package inclusions:</p> <ul style="list-style-type: none"> •Maternal Birthing Simulator, Active birthing baby, Tablet PC with rugged case, Simulator Control Software with Lifetime License, Vitals Patient Monitor, (9) Labor & Delivery Simulation Learning Experience, Automatic Mode control license, Integrated automatic drug recognition, 20 Programmable drug recognition syringes with labels, Includes preprogrammed scenarios, Streaming voice headset, RF communications module, Bluetooth communications module, USB wireless router, Mother and baby battery chargers, User guide <p>Package inclusions:</p> <ul style="list-style-type: none"> •Palpation abdominal cover, Contraction abdominal cover, C-section abdominal cover, Postpartum abdominal cover, 2 umbilical cords, 2 precut umbilical cords, Placenta, Epidural insert, C-section baby, 2 C-section abdominal skins, 2 C-Section uterine walls, Healthy PPH perineum, Midline episiotomy suture trainer, Mediolateral episiotomy suture trainer, 4th-degree episiotomy suture trainer, 2 suppositories, Artificial blood concentrate, Mineral oil lubricant, Hospital gown, baby head cap, receiving blanket and diaper, Manual blood pressure cuff, 2 baby delivery rings, Birthing baby cradle, Bladder, IV, and PPH filling kits, NIBP calibration kit, Antecubital vein replacement set, Spare cervix and birth canal <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . Training 2-3 days.</p>	
3.	<p>Maternal Fetal Birthing Simulator (HIGH FEDELITY MANIKIN)</p> <p>General Specifications: Fully Instrumented Mother and a Partially Instrumented Fetus, Muse Operating Software, MacBook Pro Instructor's Wireless Workstation, Touch-Pro Wireless Patient Monitor Computer, 10 Preconfigured SCEs, 4 SCE Development Licenses, First Year Full System Value Service Agreement Free,</p>	

<p>One Year of Training for end-users. Includes: Female Patient Module: turns your Birthing Simulator to a regular female patient simulator. Includes 5 Pre-Configured SCEs and a Non Gravid Abdomen Manikin/Simulator:</p> <ul style="list-style-type: none"> • Birthing Simulator no more than 176 cm long. • No more than 50 kg in weight • Fetus: Max 48 cm long and Max 2.5 kg in weight • Wireless patient monitor via All-in-One Touchscreen Computer • Must have Birthing Fetus, Placenta and Umbilical Cord • Separate Leopold's Maneuver Fetus • Prepartum Abdomen • Set of Static Cervices for vaginal exams • Postpartum Boggy/Contracted uterus with blood tank • Blood tank minimum of 1.8 liters to simulate Class III hemorrhage • Ability to Simulate Prepartum and Postpartum Scenarios • w/ Fully and partially inverted uterus, Uterine Inversion • Abdominal Cover for all fours birthing, Non-gravid abdomen • Upgradeable for Augmented Reality Training <p>Simulator Features: Breathing:</p> <ul style="list-style-type: none"> • Lung auscultation: anterior and posterior with individual lung control <p>Spontaneous breathing</p> <ul style="list-style-type: none"> • Spontaneous chest excursion synchronized with ventilation • Lung sounds can be independently controlled. Airway: Anatomically realistic upper airway. • Lungs allow for mechanical positive pressure ventilation support and allows ventilation modes (SIMV and CMV) • Airway Management and Ventilation, Bag-Valve-Mask • Articulated Mandible for jaw thrust maneuver • Detects right mainstem intubation • ET Tube, nasal-pharyngeal and oropharyngeal airways • Must feature an articulating mandible <p>Blinking eyes that react to light</p> <p>Light reactive and blinking can be controlled via software options</p> <ul style="list-style-type: none"> • Seizures via rhythmic movement of arms and rapid blinking • Eyes must be able to reflect the following conditions: Jaundice, Hemorrhage, Keyhole Pupil Cataract, Bloodshot, Droopy Eyelid <p>Urological:</p> <ul style="list-style-type: none"> • Urine output <p>Sounds:</p> <ul style="list-style-type: none"> • Heart, bowel and breath sounds, Vocal Sounds • Other sounds can be provided via wireless microphone • Two-way communication • Audible breathing sounds • Vocal sounds can be synchronized with uterine activity <p>Vascular Access:</p> <ul style="list-style-type: none"> • IV insertion supported in right arm: cephalic, basilic and antebrachial veins • IO site on anterior tibia • Right jugular IV line for infusions <p>Articulation:</p> <ul style="list-style-type: none"> • Range of motion in wrists, elbows, hips, knees and ankles. <p>Obstetrical:</p> <ul style="list-style-type: none"> • Integrated modeling of maternal-fetal physiology measurements 	
--	--

<ul style="list-style-type: none"> • Maternal aesthetics build from real patient Realistic birth canal and vulva/perinium supports accurate fetal descent and rotation • Birthing Positions: lithotomy, sitting and all fours • Rectal suppository administration Prepartum: • Vaginal examinations of cervix, fetal station and position • Static cervixes representing various stages of dilation, from closed to 5cm, effacement from 0 to 90% • Leopold's maneuver can be performed to determine presentation and position of fetus • Epidural port Intrapartum: • Realistic and palpable uterine contractions detected by palpating the fundus, Rate and duration of contractions can be controlled via software, Trendelenburg position with detection, Left lateral tilt with detection, Vertex and Breech delivery, Application of McRoberts Maneuver with observable pelvic tilt, supported and logged, Suprapubic pressure support and detection with palpable symphysis pubis, Supports delivery of posterior arm for shoulder dystocia • Zavanelli maneuver with detection • C-section team training support • Rotation of anterior and posterior shoulder detected and resolves shoulder dystocia when performed correctly via Rubin II and Woods' Screw Maneuvers • Forceps application • Vacuum extraction without need for fetal cap • Fetal hear sounds: 5 locations depending upon fetal presentation • Placenta: intact/fragmented with realistic color/texture/42lexibility Placenta delivery Fetus: • Birthing fetus without unrealistic connection ports in the head and/or buttocks • Articulated fetal body neck, shoulders, elbows, hips and knees • Fetus has lateral neck movement • Clinically accurate fetal size at 5th percentile of WHO growth chart All scenarios provided must be evidence based and have the following: o Synopsis <ul style="list-style-type: none"> • Scenario Programming of Patient History • SBAR Handoff Report • Healthcare Provider Orders Learning Objectives • Learning Performance Measures • Preparatory Questions • Suggested Equipment and Supplies needed of Facilitator Notes o Debriefing Points o Teaching Q & A o Clinical References (Evidence Based) Included Scenarios: 15 Scenarios Birthing Scenarios 1. Normal delivery 2. Instrumented vaginal delivery 3. Fetal Tachycardia due to maternal pyrexia 4. Breech Delivery 	
---	--

	<p>5. Fetal Central Nervous System Depression by narcotic given to mother</p> <p>6. Shoulder Dystocia</p> <p>7. Major Post-partum Hemorrhage due to uterine atony</p> <p>8. Maternal Cardio-respiratory Arrest</p> <p>9. Eclampsia</p> <p>10. Umbilical Cord Prolapse</p> <p>Non-Birthing Scenarios</p> <p>8. Chronic Heart Failure Exacerbation</p> <p>12 Acute Respiratory Distress Syndrome</p> <p>13.Sepsis with Hypotension</p> <p>14. Brain Attack w/ Thrombolytic Therapy</p> <p>15.Motor Vehicle Collision with Hypovolemic Shock</p> <p>Warranty: 2 years on parts and service.</p> <p>Preventive Maintenance: once a year for 5 years .</p> <p>Training 2-3 days.</p>	
4.	<p>Adult Training Manikin/ Geriatric Manikin</p> <p>Lightweight – approximately 28 lbs.</p> <ul style="list-style-type: none"> •Full-size adult manikin – measures 58 in. •Overall female appearance •Converts to male with removal of wig and attachment of male genitals <p>Visual Inspections:</p> <ul style="list-style-type: none"> •Normal and Cancerous Mole •Stage 1 Sacral Ulcer •Dilated and Constricted Pupils •Reddened Skin Folds <p>Patient Care Simulation:</p> <ul style="list-style-type: none"> •Bandaging and Wound Dressing •Bed Baths •Clothing Changes Patient Care Simulation: •Bandaging and Wound Dressing •Bed Baths •Clothing Changes •Denture Placement and Removal, Upper and Lower •Ear Canal Irrigation, Optic Drops, and Hearing Aid Placement •Eye Irrigation •Finger and Toe Manipulation with Bandaging – Flexible and Individually Molded •Hair Care Washing and Combing •Intramuscular Injection Sites: Arm, Thigh, Buttock •Oral and Nasal Hygiene: Lavage, Gavage, and Suctioning •Ostomy Care: Ileostomy and Colostomy Tissue •Maintenance and Appliance Application, Lavage, and Suctioning •Patient Positioning •Patient Transfer Techniques •Pericare •Tracheostomy Care, Lavage and Suctioning •Gastrostomy Procedures, Lavage and Gavage •Ankle – Eversion, Inversion, Dorsiflexion, Plantarflexion •Elbow – Extension, Flexion, Pronation, Supination •Fingers – Abduction, Adduction, Flexion (soft, lifelike material) •Hip – Abduction, Adduction, Rotation, Hyperextension •Knee – Extension, Flexion•Enema Administration, Female Only •Urinary Catheterization, Female and Uncircumcised •Neck – Rotation, Hyperextension, Lateral Flexion •Shoulder – Abduction, Adduction, Rotation, Hyperextension •Toes – Abduction, Adduction, Flexion (soft, lifelike material) 	

	<ul style="list-style-type: none"> •Trunk – Rotation, Hyperextension •Wrist – Flexion, Hyperextension, Radial Flexion, Ulnar Flexion <p>Realistic Patient Positioning, Flexibility, & Superior Range of Motion Male</p> <ul style="list-style-type: none"> • Pap Smears and Douching • Prostate Exam, Stage B • Nasogastric Tube Placement • Internal Fluid Reservoirs <p>Warranty: 2 years on parts and service. Preventive Maintenance: once a year for 5 years Training 2-3 days.</p>	
5.	<p>Human Whole Body Plastinated Specimen</p> <p>Dry & odorless, No oil, water, molds or any leaks present in the specimen to show signs of decomposition</p> <p>Left Half showing superficial muscles and other right half in deep muscle dissection. Front and back. Synopsis: Skeletal muscle is an important portion of locomotive system and is also the emphases of teaching. After delicately dissected, the specimen displays the skeletal muscle of the whole human body and some cutaneous nerves and superficial arteries which are remained simultaneously. One side of the Specimen displays superficial layers muscles and the other side displays Deep layer muscles. By means of the contrast between the two sides, students can deeply understand and grasp the position and relationship of the muscles.</p> <p>Warranty: 2 years on parts and service. Preventive Maintenance: once a year for 5 years . “Terms of reference for plastinated specimens”</p> <ol style="list-style-type: none"> 1. Plastinated specimens are preserved using specialized silicone or compound for whole body or regional parts. 2. Plastinates should be cleanly dissected to reveal specific anatomical parts and not twisted, deformed or distorted. 3. Plastinates should have a life-like natural color. 4. Plastinates should be completely dry to touch not gluey, sticky or watery. 5. Plastinates should be odorless and should not emit any odor/ foul smell or when stored in glass containers 6. Plastinates should be non toxic and does not possess any allergic reactions to the eyes, nose, skin etc. 7. Plastinates should be portable i.e not heavy 8. Distributor should replace the plastinated specimen when it does not conform to the conditions above. 9. Plastinates should have a 3-dimensional feature 10. Plastinates can be stored in free air (not airconditioned) indefinitely. Humidity ranges below 60% and can withstand temp range of 20-30 degc 11. Distributor should render free aftersales and preventive maintenance service for at least 5 years. 12. Distributor should be able to provide repair when specimens are damaged. 13. The distributor must show an actual demo unit during post qualification 	

6.	<p>Advance Patient Training Arm Subcutaneous injection sites on the volar forearm and lateral upper arm Intramuscular injection site on the upper arm Arterial system including the radial and brachial arteries Suture and incision sites on both the upper arm and forearm Hemodialysis site on the forearm Arterial and Venous insert for IV training, venipuncture practice, blood draw exercises, AV anastomosis, and placement of AV grafts. This multi-layer surgical insert includes the skin, subcutaneous tissue, muscle, radial artery, and radial vein AV fistula insert that simulates a healed fistula for hemodialysis exercises. Multi-layer bicep insert that includes the skin, subcutaneous tissue, and muscle and allows incision practice and suturing exercises. Durable skin that can be pierced in excess of 200 times with a 20 or 22 gauge needle. Realistic tactile feedback for both surgical and arterial & venous stick exercises. Adjustable heart rate and pulse strength simulating a heart rate from 10 BPM to 150 BPM. Cephalic (antecubital), Basilic, Radial, and Ulnar veins, as well as the radial and brachial arteries for infusion and blood draw training. Rotating arm allowing dorsal and volar access along the length of the IV training arm. Varying vessel palpability to simulate collapsed or bulging vessels. Ease of assembly. Latex-Free New proprietary materials replicate the skin, subcutaneous, and muscle layers at all surgical sites. Upgraded inserts, skins, and vessels for improved tactile feedback Micropump embedded within the shoulder generates variable heart rates and pulse strength. Quiet and compact, thereby improving the portability of the unit. Easy to replace plug-and-play inserts to allow quick change-out between procedures. Latex-free vessels with improved access for hassle-free replacement</p>	
7.	<p>CPR Manikins (set of 3 manikins-Adult, Child & Infant) Features: <ul style="list-style-type: none"> • CPR feedback monitor that meets 2015 AHA CPR guidelines for Adult Manikin. • Available in Adult, Child, and Infant models. • Realistic looking and performing manikin is more acceptable to both instructors and students. • Far easier and faster to carry and set up at training sites. • Patented face/head tilt simulates how an actual victim’s head would move if he required CPR and teaches students how to properly open the airway. • Easily inserted face-shield/lung-bag provides visual chest rise to ensure students receive the most accurate and realistic training. The face shield is secured so it won’t slip and there’s no need for adhesives. Prestan Manikins are not made with natural rubber latex 2 Extra Compression Springs (Hard and Soft) Training Pads, Airway Lubricant, Eye Set, Kneel Mat, Jacket, Pants Manikin Wipes, USB Cable Plus a ShockLink with Defibrillator Training Pads for ShockLink ShockLink Remote Control ShockLink Pouch and User Guide Includes 1 x Adult Manikin with CPR monitor</p>	

	<p>1 x Child Manikin with CPR monitor 1 x Infant Manikin with CPR monitor 10 x Adult Face Shield Lung Bags 10 x Child Face Shield Lung Bags 10 x Infant Face Shield Lung Bags 1 x Nylon Carrying case Warranty: 1 year</p>	
LOT 2 – HOSPITAL EQUIPMENT		
8.	<p>BABY BASSINET with stand single Bassinet stand with acrylic glass basket Bassinet stand dimension: 83 x 46 x 64.5cm Frame leg and support stainless steel Shelf-stainless steel Mattress of the bassinet is ordinary foam with cloth cover With four lockable caster wheels at 4 inches in size Basket dimension: (H) 22cm x (L) 75.5cm x (W) 40cm Warranty: 1 year</p>	
9.	<p>EMERGENCY BED STRECHER Operated by Handle. When you turn it right, it is up, turn left, it is down. The Central-Lock System and Swivel Castor can help the User easy to control and stop. Wheel Size: Φ150mm Double Sided. High position: 192x61x83cm Low position: 192x63x54cm Size of the castors: Dia.150mm Max angle of the back: 65 0 Self-weight: 85kgs Load bearing: 159kgs Packing Quantity: 1 Size(LxWxH): 198x73x64cm Warranty: 1 year</p>	
10.	<p>Medical Emergency Crash Cart 1. Mainly of aluminum, steel and ABS engineering plastic; bear Four columns to bear weight; 2. Upper part: ABS integrated board molding by injection one-time, with handralls on both sides, small items won' t slip; With transparent soft glass on the table effectively prevents scratches 3. Front: Push and pull smoothly, the Centralized Lock, Key can be folded to ensure the safety of the items inside the drawer; The drawer handle is dovetail shaped, which can effectively prevent dust from entering. Five drawers, which are divided into two small, two medium, and one large drawers. Each drawer inside with 3*3 dividers, can be separated freely. Label card holder is slot type design, to prevent liquid and dust into holder. The label card holder is inclined upward for convenient looking, which is according to the ergonomic principle. 4. Left side: Difillatoar shelf, which is used to store the defibrillator. It has fixed magic straps and does not fall off. Pull-out sliding writing frame improves space application efficiency Movable File Container enables easier access to patient information 5. Right side: Equipped with lifting infusion stand, convenient and easy to operate; One 2L needle disposal holder in the utility basket, Double ABS dust basket for waste sorting will effectively eliminate the smell of garbage; 6. Back: -CPR Board, for acute stroke patients</p>	

	<p>- Oxygen Tank Holder with belt attached to the body of the trolley. Does not drop when moving. -power outlet & hook with 5m wire. 7. Bottom: -luxury universal insert noiseless casters, two with brake, casters are made of high-strength polyurethane. -Anti- static, anti-hair winding, portable and flexible; 750*475*930mm 5 Layers Packing Size: 760*540*920 mm Rack of multi bin container Disposable lock with apple-shape, IV Pole, Defibrillator shelf Warranty: 1 year</p>	
11.	<p>INFUSION PUMP (1) and SYRINGE PUMP(1) Extendable from 3 up to 6 or 9 Pump Slots • by simply connecting 2 or 3 Standard Racks (TE*RS700N) or by adding 1 or 2 Rack Extension Devices to the Communication Rack Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of pump & sets* Stable infusion thanks to dedicated set Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Luer-Lock syringes Infusion Sets Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of Terumo pump & sets* Stable infusion thanks to dedicated set * Flow rate accuracy: within $\pm 5\%$ Available in different sizes Transparent barrel Bold and easy legible scale markings No Latex & PVC components Silicon coating Double contact gasket Infusion Sets Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of pump & sets* Stable infusion t* Flow rate accuracy: within $\pm 5\%$ Available in different sizes Transparent barrel Bold and easy legible scale markings No Latex & PVC components Silicon coating Double contact gasket Warranty: 1 year</p>	
12.	<p>INFANT DIGITAL WEIGHING SCALE Special feature Backlit Display Material: Plastic Item weight: 8 Pounds Small Footprint. Accuracy to 1 2 ounce. Detachable tray for easy cleaning and portability. Ergonomically designed for both ease of use and infant safety. features: metric lockout, integral measuring tape (up to 22 inch/56 cm), 30 lb x 0.1 oz / 15 kg x 0.005 kg capacity, locking pin for easily removing the tray for cleaning, 1 inch (25 mm) high LCD, and RS232 serial output for connectivity</p>	

	<p>with a printer or PC. The spacious weighing tray measures 26 in W x 12 in D / 66 cm W x 31 cm D. The scale's ZERO button removes blanket tare weights. Your choice of rechargeable NiCad batteries (not included), standard Alkaline batteries (not included), or UL-listed AC adapter (included). The 6745 pairs perfectly with the Detecto stainless steel baby scale cart.</p> <p>Warranty: 1 year</p>	
13.	<p>12 LEAD ECG MACHINE Channels Record & print 12 channels of data at a time, fastest speed Standard 12 leads; Cabrera; Nehb, VCG 9 Inches, 15° angled display for clear visualization Touchable, time-saving operation quick input of patient data Rechargeable, 3-hour continuous work Local Storage: 3,500 files Thermal Paper:Z-fold paper: 210*140mm Economic Recording Mode Weight: 3.4KG Accessories in the Package ECG Acquisition Box with Lead Wire Z-fold Thermal Paper (160 Pieces: 112mm × 140mm) Power Cord ×2 (US & EU Standard) & AC Power Adapter Equipotential Grounding Cable Warranty: 1 year</p>	
14.	<p>Orthopedic bed with balkan frame, 3 cranks ·Firm mattress ·Fracture board ·Bed elevator/ shock block (optional) ·Balkan frame (cross bar, curve bar, vertical, horizontal, diagonal bars) ·3 pulleys ·3 clamps ·Overhead trapeze Dimensions: L2240xW1060xH470=680mm Function : 3 functions Material : Cold rolled steel, HDPE, Aluminum alloy Side rails : ABS Side Rails Material: Cold rolled steel, HDPE, Aluminum alloy Dimensions : L2240xW1060xH470=680mm Bed platform : 1980x900mm Backrest : 0-75deg +-5deg Legrest : 0-45deg +-5deg Height adjustment : 470-680+-10mm Brake system : Central brake Sticker color : Wood Grain/Lake Blue Side rails : ABS side rails Orthopedics traction shelf : Aluminum Alloy 2023”793”1503mm Features 1. Blue liquid angle indicator on backrest and leggiest side rails 2. Clear acrylic head card for patient name display, four infusion holes, four drainage hooks around the bed 3. Mechanical crank operation with overload protection 4. The screw handle is foldable under the bed footboard 5. Four 125mm diameter castors with central brake 6. Orthopedics traction is made of molding aluminum alloy 2023*793*1503mm Warranty: 1 year</p>	

15.	<p>Balance Skeletal Traction Equipment</p> <ul style="list-style-type: none"> · Thomas splint (with half ring) · Pearson's attachment · Rest splint · Steinmann's pin holder · Braun-Bohler splint · Slings with clips or pins · 3 sash cords (thigh rope, traction rope, suspension rope) · 2 weight bags <p>Traction weight (10% of patient's body weight) Suspension weight (1/2 of traction weight)</p> <ul style="list-style-type: none"> · Foot pedal or foot board <p>Warranty: 1 year</p>	
16.	<p>Limb (Foot) use for Balance Skeletal Traction (1 adult and 1 Pedia Limb Warranty: 1 year</p>	
17.	<p>OR LIGHT SINGLE HEAD</p> <p>max illumination power 160,000lux diameter of lighthouse 670mm color temperature 3000k-5000k color rendering index 96% dimming range 1-100% light field diameter d10 24-36cm/9.5"-14.2" light field diameter d50 12-28cm/5.1"-11" number of leds 60leds lifetime of led >60,000hours degree of turbulence (standard value 37.5% max) 29% depth of volume of light L1 + L2 at 20% 130cm/51.2" depth of volume of light L1 + L2 at 60% 76cm/30" irradiance at nominal illuminance 510w/m² handle holder made of ABS plastic flexible arms that provide comfortable positioning there is a laser marking feature to mark the midpoint of the central illumination intensity endo light mode/standby mode can be switch by a single button Warranty: 1 year</p>	
18.	<p>OPERATING ROOM TABLE</p> <p>Electro Hydraulic OR Table Easy to handle modular design for different surgical disciplines Long and strong rail line for accessories T base design for maximum approach Hand control and column control units Battery back up: up to 30 full movements</p> <p>Technical Features: Sliding Movement: 300mm Lateral Tilt Movement: +20 degree Celsius / -20 degree Celsius Back Rest Movement: +80 degree Celsius / -45 degree Celsius Trendelenburg: +30 degree Celsius / -80 degree Celsius Reverse Trendelenburg: -30 degree Celsius / -90 degree Celsius Leg Rest Up/Down: +30 degree Celsius / -80 degree Celsius Head Rest Movement: +30 degree Celsius / -90 degree Celsius Flex/Reflex Position: +225 degree Celsius / -100 degree Celsius Width x Length: 560mm x 2200mm Height (Min): 595mm Height (Max): 1035mm Mattress Thickness: 70mm Castor Diameter: 100mm</p>	

	<p>Maximum Patient Weight: 360kg Stainless Steel T Base Geopel Leg Holder -made of high-quality stainless steel -pair, including clamp -height adjustable -rotatable and tillable via ball-join Warranty: 1 year</p>	
19.	<p>Digital Binocular Biological Microscope Optical System: Infinite Optical System Eyepiece: WF10x / 20mm, ANTI FUNGUS with pointer Viewing Head: Seidentopf Binocular Head, inclined at 30°, Interpupillary 47-78mm ANTI FUNGUS Objective: Infinite Semi-Plan Achromatic Objectives 4×, 10×, 40×, 100× ANTI FUNGUS Nosepiece: Backward Quadruple Nosepiece Stage: RACKLESS Double Layers Mechanical Stage 150mm×139mm, Moving Range 75mm×52mm Condenser: Abbe Condenser NA1.25 Focusing: Coaxial Coarse and Fine Adjustment, Fine Division 20mm, Moving Range 0.002mm Illumination: S-LED Illumination with Rechargeable Battery, Brightness Adjustable, can be connected to an LCD monitor or HD TV via HDMI cable and operated independently without connecting to PC. The image/video capture and operate can be controlled by mouse, so no shaking when you take images and videos. It also can be connected to a PC via USB2.0 cable and operate with the software. Features Use mouse to control the Camera. Record image and video to SD card. High frame rate of 15fps. The functions inside camera (Cloud 1.0)The less Icons the Better. Set Exposure Time Capability. 3D Noise Reduction. 1080P Video Recording. Get more Details with ROI Magnification Function. Image Comparison Function. Browse Captured Images PC software Warranty: 1 year</p>	
20.	<p>Electric Wheelchair Total length of wheelchair: 920MM Total height of wheelchair body: 850MM Seat width: 460MM Rear wheel size: 12 inch Front wheel size: 8 inch Frame material: steel Motor power: 250W (2) 60HZ Climbing capacity: 35 ° Net weight 30KG Both sides open and close armrest Breathable cushion High-strength ABS anti-lashing rear wheel Thickened footrest Thickened steel</p>	

	Adjustable seat belt Back-up reminder Large capacity back pocket Anti-dumping rear wheel EPBS braking system without skidding Warranty: 1 year	
21.	WEIGHING SCALE WTH HEIGHT MEASUREMENT Heavy-duty solid stable platform Dual-reading die-cast weight beam which may be read from either side of the scale Durable steel construction with an electrostatic powder paint finish for optimum quality. Capacity may be increased by using an optional counterweight Capacity: 400 lb x 4 oz , 180 kg x 100 g, 400 lb x 4 oz / 175 kg x 100 g Scale Height: 59 in / 150 cm Height Rod (on select models - 2392 excluded): 30 to 78 in / 76 to 200 cm - 1/4 in / 1/2 cm increments Platform Size: 10.5 in W x 14.5 in D / 27 cm W x 37 cm Warranty: 1 year	
22.	HOSPITAL BED Electric multi-functional nursing bed medical bed hospital bed turn over bed rehabilitation lift bed Intelligent remote control operation (also supports manual) Sending shampoo basin, universal wheel, urinal, infusion rack, dining board International certified motor (38DB) low noise Can be timed to turn over (30-45 minutes once) Electric start back One piece reset Electric lift leg (0-25°) Electric drop leg (0-85°) Electric turning (0-45°) Electric helper for defecation Head washing function Foot washing function Infusion function Meal function Mobility function Automatic lifting of the guardrail Anti-wetting mattress Size:2060*1030*480mm Bed size:1950*900mm. Warranty: 1 year	
23.	36" Discover Headwall Package PRODUCT FEATURES •The Functioning Wall Mounted 36" Discover Headwall Package is the most popular training headwall. It simulates oxygen, air and suction with one source of air by connecting to an external Silent Air Compressor. The Discover headwall package is designed for simulation labs so it is easy to install and maintain, and is available in a variety of sizes and color options, Discover size: 39.38" H x 19.5" W x 4.75" D, Wall mounted: Mounting strip included, Nurse call facade with code blue button and hand pendant TECHNICAL DATA • Size: 36", Wall Mounted – Includes mounting strip, Ohmeda gas outlets, Oxygen, Air, Vacuum, Duplex receptacle, Vacuum slide, Nurse call 3 gang cover, Includes Complete Accessories Package	

	•Oxygen Flow Meter, Medical Air Flow Meter, Analog Vacuum Regulator, Suction Canister, Humidifiers, Nasal Cannulas, Suction tubing, DISS Hose and Fittings Warranty: 1 year	
--	--	--

Attached herewith are the manufacturer's product literature(s) and certification(s) that we are authorized to sell the goods.

We certify that the foregoing information and the supporting documents are true and correct.

[Signature]
[Name of Authorized Signatory]
[Position/Title of Authorized Signatory]

Omnibus Sworn Statement

REPUBLIC OF THE PHILIPPINES)
CITY/MUNICIPALITY OF _____) S.S.

AFFIDAVIT

I, [Name of Affiant], of legal age, [Civil Status], [Nationality], and residing at [Address of Affiant], after having been duly sworn in accordance with law, do hereby depose and state that:

1. **Select one, delete the other:**

- b. *If a sole proprietorship:* I am the sole proprietor or authorized representative of [Name of Bidder] with office address at [address of Bidder];
- c. *If a partnership, corporation, cooperative, or joint venture:* I am the duly authorized and designated representative of [Name of Bidder] with office address at [address of Bidder];

2. **Select one, delete the other:**

- a. *If a sole proprietorship:* As the owner and sole proprietor, or authorized representative of [Name of Bidder], I have full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached duly notarized Special Power of Attorney;
 - b. *If a partnership, corporation, cooperative, or joint venture:* I am granted full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached [state title of attached document showing proof of authorization (e.g., duly notarized Secretary's Certificate, Board/Partnership Resolution, or Special Power of Attorney, whichever is applicable)];
3. [Name of Bidder] is not "blacklisted" or barred from bidding by the Government of the Philippines or any of its agencies, offices, corporations, or Local Government Units, foreign government/foreign or international financing institution whose blacklisting rules have been recognized by the Government Procurement Policy Board, by itself or by relation, membership, association, affiliation, or controlling interest with another blacklisted person or entity as defined and provided for in the Uniform Guidelines on Blacklisting;
4. Each of the documents submitted in satisfaction of the bidding requirements is an authentic copy of the original, complete, and all statements and information provided therein are true and correct;
5. [Name of Bidder] is authorizing the Head of the Procuring Entity or its duly authorized representative(s) to verify all the documents submitted;

6. **Select one, delete the rest:**

- a. *If a sole proprietorship:* The owner or sole proprietor is not related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;
- b. *If a partnership or cooperative:* None of the officers and members of [Name of Bidder] is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of

the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

- c. *If a corporation or joint venture:* None of the officers, directors, and controlling stockholders of *[Name of Bidder]* is related to the Head of the Procuring Entity, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

7. *[Name of Bidder]* complies with existing labor laws and standards; and

8. *[Name of Bidder]* is aware of and has undertaken the following responsibilities as a Bidder:

- a. Carefully examine all of the Bidding Documents;
- b. Acknowledge all conditions, local or otherwise, affecting the implementation of the Contract;
- c. Made an estimate of the facilities available and needed for the contract to be bid, if any; and
- d. Inquire or secure Supplemental/Bid Bulletin(s) issued for the *[Name of the Project]*.

9. *[Name of Bidder]* did not give or pay directly or indirectly, any commission, amount, fee, or any form of consideration, pecuniary or otherwise, to any person or official, personnel or representative of the government in relation to any procurement project or activity.

10. In case advance payment was made or given, failure to perform or deliver any of obligations and undertakings in the contract shall be sufficient grounds to constitute criminal liability for Swindling (Estafa) or the commission of fraud with unfaithfulness or abuse of confidence through misappropriating or converting any payment received by a person or entity under an obligation involving the duty to delivery certain goods or services, to the prejudice of the public and the government of the Philippines pursuant to Article 315 of Act No. 3815 s. 1930, as amended, or the Revised Penal Code.

IN WITNESS WHEREOF, I have hereunto set my hand this ___ day of ___, 20__ at _____, Philippines.

Bidder's Representative/Authorized Signatory

SUBSCRIBED AND SWORN to before me this ___ day of *[month]* *[year]* at *[place of execution]*, Philippines. Affiant/s is/are personally known to me and was/were identified by me through competent evidence of identity as defined in the 2004 Rules on Notarial Practice (A.M. No. 02-8-13-SC). Affiant/s exhibited to me his/her *[insert type of government identification card used]*, with his/her photograph and signature appearing thereon, with no. _____ and his/her Community Tax Certificate No. _____ issued on ___ at _____.

Witness my hand and seal this ___ day of *[month]* *[year]*.

NAME OF NOTARY PUBLIC

Serial No. of Commission _____

Notary Public for _____ until _____

Roll of Attorneys No. _____

PTR No. _____ *[date issued]*, *[place issued]*

IBP No. _____ *[date issued]*, *[place issued]*

Doc. No. _____
Page No. _____
Book No. _____
Series of _____

[Bidder's Letterhead]

[Date]

To: Tarlac State University
Re: Invitation to Bid No.

NET FINANCIAL CONTRACTING CAPACITY

Based on our Income Tax Return and Audited Financial Statement for the Fiscal Year [YEAR], duly submitted to the Bureau of Internal Revenue, and which form part of our Bid, the summary of our firm's financial condition is as given below:

		Year [YEAR]
1.	Total Assets	
2.	Current Assets	
3.	Total Liabilities	
4.	Current Liabilities	
5.	Net Worth (1-3)	
6.	Net Working Capital (2-4)	

Based on the aforementioned data and the Value of Outstanding Works from the Statement of All Ongoing Government and Private Contracts, which also form part of our Bid, our Net Financial Contracting Capacity (NFCC) is:

NFCC = [(current asset minus current liabilities) (**15**)] minus [value of all outstanding or uncompleted portions of the projects under ongoing contracts including awarded contracts yet to be started coinciding with the contract to be bid].

NFCC =

I/We certify that the foregoing information and all of the supporting documents are true and correct.

[Signature]
[Name of Bidder or Authorized Representative]
[Position or Title]

[Bidder's Letterhead]

[Date]

FINANCIAL BID FORM

To: Tarlac State University
Re: Invitation to Bid No.

Having examined the Bidding Documents [*insert if any or delete, if none: including Bid Bulletin Numbers [insert numbers], the receipt of which is hereby duly acknowledged*], we, the undersigned, offer to [*supply/deliver/perform*] [*description of the Goods*] in conformity with the said Bidding Documents for the sum of [*total Bid amount in words and figures*] or such other sums as may be ascertained in accordance with the Schedule of Prices attached herewith and made part of this Bid.

If our Bid is accepted, we commit to deliver the goods in accordance with the delivery schedule specified in the Schedule of Requirements.

If our Bid is accepted, we undertake to provide a performance security in the form, amounts, and within the times specified in the Bidding Documents.

We agree to abide by this Bid for the Bid Validity Period specified in **ITB** Clause 14.2 and it shall remain binding upon us and may be accepted at any time before the expiration of that period.

Until a formal Contract is prepared and executed, this Bid, together with your written acceptance thereof and your Notice of Award, shall be binding upon us.

We understand that you are not bound to accept the Lowest Calculated Bid or any Bid you may receive.

We certify/confirm that we comply with the eligibility requirements as per **ITB** Clause 5 of the Bidding Documents.

We likewise certify/confirm that the undersigned, [*for sole proprietorships, insert: as the owner and sole proprietor or authorized representative of [Bidder], has the full power and authority to participate, submit the bid, and to sign and execute the ensuing contract, on the latter's behalf for the [Name of Project] of the Tarlac State University*] [*for partnerships, corporations, cooperatives, or joint ventures, insert: is granted full power and authority by the [Bidder], to participate, submit the bid, and to sign and execute the ensuing contract on the latter's behalf for [Name of Project] of the Tarlac State University*].

We acknowledge that failure to sign each and every page of this Financial Bid Form, including the attached Schedule of Prices, shall be a ground for the rejection of our bid.

Dated this _____ day of _____ 20_____.

[signature]

[in the capacity of]

Duly authorized to sign Bid for and on behalf of _____

[Bidder's Letterhead]

[Date]

To: Tarlac State University
 Re: Invitation to Bid No.

SCHEDULE OF PRICES

Item No.	Description	Qty.	Unit	Unit Price	Total Price
Lot No. 1 Manikin					
1.	<p>PEDIATRIC SIMULATOR (HIGH FEDELITY)</p> <p>General</p> <ul style="list-style-type: none"> • Height: 44 inches • Tetherless and wireless; fully responsive during transport • The internal rechargeable battery provides hours of tetherless operation • Smooth and supple full-body skin with seamless trunk and limb joints • Realistic joint articulation: neck, shoulder, elbow, hip, and knee • Palpable bony landmarks • Forearm pronation and supination • Supports common patient positions including Fowler's, supine, and sitting • Male/female patient conversion • Tablet PC preloaded with UNI® included • Includes 10 preprogrammed SLEs and facilitator's guidebook <p>Neurological</p> <ul style="list-style-type: none"> • Active robotics simulate lifelike facial expressions including: anger, transient pain, ongoing pain, amazement, Quizzical, crying, yawning Preprogrammed emotional states automatically express associated verbal and non-verbal cues without manual input: Worried, Anxious, Lethargic, Distracted • Abnormal eye and eyelid movements: crosseyed, nystagmus, eyelid twitching, eyelid droop, Programmable crying/ tears release real fluid, Wireless streaming voice: listen to participants respond in real-time, Real-time voice modulation effects, Automatic jaw movement synchronized with speech, Automatic jaw movement synchronized with speech, Seizures with selectable intensity levels, with prerecorded speech responses • Real CO2 exhalation: supports etCO2 monitoring using real sensors and monitoring devices, Selectable normal and abnormal sounds: upper right, front and 	1	pc		

<p>back; upper left, front and back; lower right, back; and lower left, back, uses “Real” mechanical ventilation support, Real-time ventilation feedback, Visible chests rise during BVM ventilation, Chest tube insertion: left midaxillary hemothorax site features palpable bony landmarks, realistic skin for cutting and suturing, tactile pleural pop, and fluid drain, Needle decompression site features realistic tactile feedback and audible hiss, Needle decompression and chest tube insertion</p> <p>Cardiac</p> <ul style="list-style-type: none"> •Includes comprehensive library of ECG rhythms with customizable beat variations, Independent normal/abnormal heart sounds at aortic, pulmonic, and mitral sites, Supports ECG monitoring using real devices <p>Supports ECG-derived respiration monitoring (EDR), Ecpr™ Real-time quality feedback and reporting</p> <p>Defibrillate, cardiovert and pace using “real devices” and energy Circulatory, Visible cyanosis, redness, pallor, and jaundice, Supports capillary refill time testing above the right knee; test detection and logging, Palpable pulses: bilateral carotid, brachial, radial, femoral, and pedal, Blood pressure-dependent pulses, Supports blood pressure monitoring using a “real NIBP” cuff and monitor, SpO2 monitoring using real devices</p> <p>Vascular Access, Bilateral forearm IV access supports sampling and continuous infusion, Intraosseous infusion site at right proximal tibia, Real glucose test readings via finger-stick</p> <p>Gastrointestinal, Patent esophagus, Gastric distension during excessive PPV, Bowel sounds in four quadrants, Interchangeable male/female genitalia, Supports urinary catheterization with fluid return, Programmable urinary output</p> <p>Airway</p> <ul style="list-style-type: none"> •Anatomically accurate oral cavity and airway, Supports nasotracheal/ orotracheal intubation with standard instruments including endotracheal tubes and supraglottic airway devices, Tracheal intubation detection, Head tilt, chin lift, jaw thrust, Supports esophageal intubation, NG/OG tube placement, Supports bag-valve-mask ventilation, Realistic surgical trachea permits tracheostomy, cricothyrotomy, and retrograde intubation, Programmable difficult airway: laryngospasm and tongue edema, 				
--	--	--	--	--

	<p>Selectable normal and abnormal upper airway sounds</p> <p>Breathing</p> <ul style="list-style-type: none"> •Spontaneous breathing and selectable normal and abnormal respiratory patterns •Variable respiratory rates and inspiratory/expiratory ratios •Programmable unilateral chest rise and fall •Unilateral chest rise with right mainstem intubation <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . Training 2-3 days.</p>				
2.	<p>Maternal and Neonatal Birthing Simulator (ULTRA HIGH FIDELITY MANIKIN)</p> <p>GENERAL SPECIFICATION</p> <ul style="list-style-type: none"> •Tetherless and wireless; fully responsive during transport, Wireless control at distances up to 300 ft.Internal rechargeable battery provides up to 10 hrs. of tetherless operation, Smooth and supple full-body skin with seamless articulating joints, Lifelike joint articulation: neck, shoulder, elbow, wrist, hip, knee, and ankle, Internal high-capacity fluid reservoirs, Maternal® Fetus-Newborn wireless link capability, Includes touchscreen tablet PC preloaded with software, Includes touchscreen virtual patient monitor and CTG <p>TECHNICAL DATA SHEET</p> <p>Processor : 12th Gen Intel® Core™ i5-12500 (18 MB cache, 6 cores, 12 threads, 3.00 GHz to 4.60 GHz Turbo, 65 W) Operating System : Windows 11 Pro Graphics Card : Intel® Graphics Monitor : OptiPlex All-in-One Non-Touch Panel Memory * : 8 GB, 1 x 8 GB, DDR4 Hard Drive : 256 GB, M.2 2230, PCIe NVMe, SSD, Class 35 Keyboard : Wireless Keyboard and Mouse Ports on the Side : 1 USB 3.2 Gen 2x1 Type-C® port 1 USB 3.2 Gen 1 port with PowerShare 1 Universal audio port Rear : 2 USB 3.2 Gen 2 ports 2 USB 3.2 Gen 1 ports with Smart Power On 1 RJ45 Ethernet port 1 Line-out audio port 1 DisplayPort++ 1.4a/HDCP 2.3 port 1 HDMI-IN—HDMI 1.4a port</p>	1	pc		

(with UMA configuration-65 W)
 1 HDMI-OUT—HDMI 2.0 port
 (with UMA configuration-65 W)
 Slots : 1 SD-card slot (side)
 1 M.2 2230 slot for Wi-Fi and Bluetooth
 card
 1 M.2 2230/2280 slot for SSD (35 W)
 2 M.2 2230/2280 slot for SSD (65 W)

NEUROLOGICAL

•Programmable blinking rate, pupil response, and bilateral and unilateral eye movement, Interactive eyes can follow a moving object, Seizures with selectable intensity levels, Wireless streaming voice: be the voice of Victoria and listen to participants’ responses via a wireless headset, Record and playback vocal responses in any language

AIRWAY: Head tilt/chin lift/jaw thrust, Oral or nasal intubation, Difficult airway: laryngospasm, tongue edema, and pharyngeal swelling.

BREATHING: Spontaneous breathing; selectable respiratory patterns and lung sounds, Programmable respiratory rates and inspiratory/expiratory ratios Realistic chest rise during assisted ventilation, Ventilation performance, real-time monitoring, reporting, and logging,

CARDIAC: Real-time compression and ventilation performance, Effective compressions generate palpable pulses and ECG artifacts. Real-time 4-Lead electrocardiogram monitoring using real ECG devices, Defibrillation, cardioversion, and pacing using real energy, Realistic heart sounds, Programmable heart rate synchronized with ECG and pulses, Virtual 12-lead dynamic ECG rhythms

CIRCULATION

•Bilateral palpable carotid, radial, and brachial pulses are synchronized with heart rate and blood pressure, Monitor oxygen saturation using real oximeters, Measure blood pressure using a real manual or automatic blood pressure cuff, Bilateral venous access, Automatic drug recognition detects medication type, dose, and rate injected into the lower right arm, SubQ and intramuscular injection sites for placement exercises

OBSTETRICS

•Maternal-fetal physiologic link, Supports Leopold’s Maneuvers and external cephalic version, Supports fetal monitoring using real devices, Precision Delivery System:

<p>computer-controlled fetal descent and cardinal movements, Programmable normal, breech, shoulder dystocia, instrument assisted delivery, and C-section, Software activated self-lubricating birth canal, Realistic placenta with detachable fragments and umbilical cord</p> <p>OBSTETRICS</p> <ul style="list-style-type: none"> •Epidural placement and needle detection; palpable anatomical landmarks and skin layers, Intrapartum bleeding, Force sensor reports traction and torsion applied to the fetus in real-time, McRoberts maneuver, suprapubic pressure, Zavanelli, Woods' Screw, and "hands and knees" position detection, Births full-term neonate of realistic size and weight, Active neonate: programmable heart and respiratory sounds, crying, cyanosis, and movement, Replaceable episiotomy inserts support suturing, Multi-layer fourth degree episiotomy with hemorrhaging vaginal sidewall tears and cervical lacerations, Programmable uterine firmness and hemorrhaging supports sutures and balloon tamponade, Programmable hemorrhage flow control <p>GASTROINTESTINAL</p> <ul style="list-style-type: none"> •Selectable bowel sounds •Internal fluid bladder with urethra for Foley catheterization exercises •Rectum with suppository placement sensor <p>Package inclusions:</p> <ul style="list-style-type: none"> •Maternal Birthing Simulator, Active birthing baby, Tablet PC with rugged case, Simulator Control Software with Lifetime License, Vitals Patient Monitor, (9) Labor & Delivery Simulation Learning Experience, Automatic Mode control license, Integrated automatic drug recognition, 20 Programmable drug recognition syringes with labels, Includes preprogrammed scenarios, Streaming voice headset, RF communications module, Bluetooth communications module, USB wireless router, Mother and baby battery chargers, User guide <p>Package inclusions:</p> <ul style="list-style-type: none"> •Palpation abdominal cover, Contraction abdominal cover, C-section abdominal cover, Postpartum abdominal cover, 2 umbilical cords, 2 precut umbilical cords, Placenta, Epidural insert, C-section baby, 2 C-section abdominal skins, 2 C-Section uterine walls, Healthy PPH perineum, 				
--	--	--	--	--

	<p>Midline episiotomy suture trainer, Mediolateral episiotomy suture trainer, 4th-degree episiotomy suture trainer, 2 suppositories, Artificial blood concentrate, Mineral oil lubricant, Hospital gown, baby head cap, receiving blanket and diaper, Manual blood pressure cuff, 2 baby delivery rings, Birthing baby cradle, Bladder, IV, and PPH filling kits, NIBP calibration kit, Antecubital vein replacement set, Spare cervix and birth canal</p> <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years. Training 2-3 days.</p>				
3.	<p>Maternal Fetal Birthing Simulator (HIGH FEDELITY MANIKIN) General Specifications: Fully Instrumented Mother and a Partially Instrumented Fetus, Muse Operating Software, MacBook Pro Instructor's Wireless Workstation, Touch-Pro Wireless Patient Monitor Computer, 10 Preconfigured SCEs, 4 SCE Development Licenses, First Year Full System Value Service Agreement Free, One Year of Training for end-users. Includes: Female Patient Module: turns your Birthing Simulator to a regular female patient simulator. Includes 5 Pre-Configured SCEs and a Non Gravid Abdomen Manikin/Simulator:</p> <ul style="list-style-type: none"> • Birthing Simulator no more than 176 cm long. • No more than 50 kg in weight • Fetus: Max 48 cm long and Max 2.5 kg in weight • Wireless patient monitor via All-in-One Touchscreen Computer • Must have Birthing Fetus, Placenta and Umbilical Cord • Separate Leopold's Maneuver Fetus • Prepartum Abdomen • Set of Static Cervices for vaginal exams • Postpartum Boggy/Contracted uterus with blood tank • Blood tank minimum of 1.8 liters to simulate Class III hemorrhage • Ability to Simulate Prepartum and Postpartum Scenarios • w/ Fully and partially inverted uterus, Uterine Inversion • Abdominal Cover for all fours birthing, Non-gravid abdomen • Upgradeable for Augmented Reality Training 	1	pc		

<p>Simulator Features: Breathing:</p> <ul style="list-style-type: none"> • Lung auscultation: anterior and posterior with individual lung control <p>Spontaneous breathing</p> <ul style="list-style-type: none"> • Spontaneous chest excursion synchronized with ventilation • Lung sounds can be independently controlled. <p>Airway:</p> <p>Anatomically realistic upper airway.</p> <ul style="list-style-type: none"> • Lungs allow for mechanical positive pressure ventilation support and allows ventilation modes (SIMV and CMV) • Airway Management and Ventilation, Bag-Valve-Mask • Articulated Mandible for jaw thrust maneuver • Detects right mainstem intubation • ET Tube, nasal-pharyngeal and oropharyngeal airways • Must feature an articulating mandible <p>Blinking eyes that react to light</p> <p>Light reactive and blinking can be controlled via software options</p> <ul style="list-style-type: none"> • Seizures via rhythmic movement of arms and rapid blinking • Eyes must be able to reflect the following conditions: Jaundice, Hemorrhage, Keyhole Pupil Cataract, Bloodshot, Droopy Eyelid <p>Urological:</p> <ul style="list-style-type: none"> • Urine output <p>Sounds:</p> <ul style="list-style-type: none"> • Heart, bowel and breath sounds, Vocal Sounds • Other sounds can be provided via wireless microphone • Two-way communication • Audible breathing sounds • Vocal sounds can be synchronized with uterine activity <p>Vascular Access:</p> <ul style="list-style-type: none"> • IV insertion supported in right arm: cephalic, basilic and antebrachial veins • IO site on anterior tibia • Right jugular IV line for infusions <p>Articulation:</p> <ul style="list-style-type: none"> • Range of motion in wrists, elbows, hips, knees and ankles. <p>Obstetrical:</p> <ul style="list-style-type: none"> • Integrated modeling of maternal-fetal physiology measurements • Maternal aesthetics build from real patient <p>Realistic birth canal and vulva/perinium supports accurate fetal descent and rotation</p> <ul style="list-style-type: none"> • Birthing Positions: lithotomy, sitting and all fours • Rectal suppository administration 				
---	--	--	--	--

<p>Prepartum:</p> <ul style="list-style-type: none"> • Vaginal examinations of cervix, fetal station and position • Static cervixes representing various stages of dilation, from closed to 5cm, effacement from 0 to 90% • Leopold’s maneuver can be performed to determine presentation and position of fetus • Epidural port <p>Intrapartum:</p> <ul style="list-style-type: none"> • Realistic and palpable uterine contractions detected by palpating the fundus, Rate and duration of contractions can be controlled via software, Trendelenburg position with detection, Left lateral tilt with detection, Vertex and Breech delivery, Application of McRoberts Maneuver with observable pelvic tilt, supported and logged, Suprapubic pressure support and detection with palpable symphysis pubis, Supports delivery of posterior arm for shoulder dystocia • Zavanelli maneuver with detection • C-section team training support • Rotation of anterior and posterior shoulder detected and resolves shoulder dystocia when performed correctly via Rubin II and Woods’ Screw Maneuvers • Forceps application • Vacuum extraction without need for fetal cap • Fetal hear sounds: 5 locations depending upon fetal presentation • Placenta: intact/fragmented with realistic color/texture/64lexibility Placenta delivery <p>Fetus:</p> <ul style="list-style-type: none"> • Birthing fetus without unrealistic connection ports in the head and/or buttocks • Articulated fetal body neck, shoulders, elbows, hips and knees • Fetus has lateral neck movement • Clinically accurate fetal size at 5th percentile of WHO growth chart All scenarios provided must be evidence based and have the following: <p>o Synopsis</p> <ul style="list-style-type: none"> • Scenario Programming of Patient History • SBAR Handoff Report • Healthcare Provider Orders Learning Objectives • Learning Performance Measures • Preparatory Questions • Suggested Equipment and Supplies needed of 				
---	--	--	--	--

	<p>Facilitator Notes</p> <ul style="list-style-type: none"> o Debriefing Points o Teaching Q & A o Clinical References (Evidence Based) <p>Included Scenarios: 15 Scenarios</p> <p>Birth Scenarios</p> <ol style="list-style-type: none"> 1. Normal delivery 2. Instrumented vaginal delivery 3. Fetal Tachycardia due to maternal pyrexia 4. Breech Delivery 5. Fetal Central Nervous System Depression by narcotic given to mother 6. Shoulder Dystocia 7. Major Post-partum Hemorrhage due to uterine atony 8. Maternal Cardio-respiratory Arrest 9. Eclampsia 10. Umbilical Cord Prolapse <p>Non-Birthing Scenarios</p> <ol style="list-style-type: none"> 9. Chronic Heart Failure Exacerbation 12 Acute Respiratory Distress Syndrome 13.Sepsis with Hypotension 14. Brain Attack w/ Thrombolytic Therapy 15.Motor Vehicle Collision with Hypovolemic Shock <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years Training 2-3 days.</p>				
4.	<p>Adult Training Manikin/ Geriatric Manikin</p> <p>Lightweight – approximately 28 lbs.</p> <ul style="list-style-type: none"> •Full-size adult manikin – measures 58 in. •Overall female appearance •Converts to male with removal of wig and attachment of male genitals <p>Visual Inspections:</p> <ul style="list-style-type: none"> •Normal and Cancerous Mole •Stage 1 Sacral Ulcer •Dilated and Constricted Pupils •Reddened Skin Folds <p>Patient Care Simulation:</p> <ul style="list-style-type: none"> •Bandaging and Wound Dressing •Bed Baths •Clothing Changes Patient Care Simulation: <p>Simulation:</p> <ul style="list-style-type: none"> •Bandaging and Wound Dressing •Bed Baths •Clothing Changes •Denture Placement and Removal, Upper and Lower •Ear Canal Irrigation, Optic Drops, and Hearing Aid Placement •Eye Irrigation •Finger and Toe Manipulation with Bandaging – Flexible and Individually Molded 	4	pc		

	<ul style="list-style-type: none"> •Hair Care Washing and Combing •Intramuscular Injection Sites: Arm, Thigh, Buttock •Oral and Nasal Hygiene: Lavage, Gavage, and Suctioning •Ostomy Care: Ileostomy and Colostomy Tissue •Maintenance and Appliance Application, Lavage, and Suctioning •Patient Positioning •Patient Transfer Techniques •Pericare •Tracheostomy Care, Lavage and Suctioning •Gastrostomy Procedures, Lavage and Gavage •Ankle – Eversion, Inversion, Dorsiflexion, Plantarflexion •Elbow – Extension, Flexion, Pronation, Supination •Fingers – Abduction, Adduction, Flexion (soft, lifelike material) •Hip – Abduction, Adduction, Rotation, Hyperextension •Knee – Extension, Flexion •Enema Administration, Female Only •Urinary Catheterization, Female and Uncircumcised •Neck – Rotation, Hyperextension, Lateral Flexion •Shoulder – Abduction, Adduction, Rotation, Hyperextension •Toes – Abduction, Adduction, Flexion (soft, lifelike material) •Trunk – Rotation, Hyperextension •Wrist – Flexion, Hyperextension, Radial Flexion, Ulnar Flexion <p>Realistic Patient Positioning, Flexibility, & Superior Range of Motion Male</p> <ul style="list-style-type: none"> • Pap Smears and Douching • Prostate Exam, Stage B • Nasogastric Tube Placement • Internal Fluid Reservoirs <p>Warranty: 2 years on parts and service. Preventive Maintenance: once a year for 5 years . Training 2-3 days.</p>				
5.	<p>Human Whole Body Plastinated Specimen Dry & odorless, No oil, water, molds or any leaks present in the specimen to show signs of decomposition Left Half showing superficial muscles and other right half in deep muscle dissection. Front and back. Synopsis: Skeletal muscle is an important portion of locomotive system and is also the emphases of</p>	1	pc		

	<p>teaching. After delicately dissected, the specimen displays the skeletal muscle of the whole human body and some cutaneous nerves and superficial arteries which are remained simultaneously. One side of the Specimen displays superficial layers muscles and the other side displays Deep layer muscles. By means of the contrast between the two sides, students can deeply understand and grasp the position and relationship of the muscles.</p> <p>Warranty: 2 years on parts and service. Preventive Maintenance:- once a year for 5 years . “Terms of reference for plastinated specimens”</p> <ol style="list-style-type: none"> 1. Plastinated specimens are preserved using specialized silicone or compound for whole body or regional parts. 2. Plastinates should be cleanly dissected to reveal specific anatomical parts and not twisted, deformed or distorted. 3. Plastinates should have a life-like natural color. 4. Plastinates should be completely dry to touch not gluey, sticky or watery. 5. Plastinates should be odorless and should not emit any odor/ foul smell or when stored in glass containers 6. Plastinates should be non toxic and does not possess any allergic reactions to the eyes, nose, skin etc. 7. Plastinates should be portable i.e not heavy 8. Distributor should replace the plastinated specimen when it does not conform to the conditions above. 9. Plastinates should have a 3-dimensional feature 10. Plastinates can be stored in free air (not airconditioned) indefinitely. Humidity ranges below 60% and can withstand temp range of 20-30 degc 11. Distributor should render free aftersales and preventive maintenance service for at least 5 years. 12. Distributor should be able to provide repair when specimens are damaged. 13. The distributor must show an actual demo unit during post qualification 				
6.	<p>Advance Patient Training Arm Subcutaneous injection sites on the volar forearm and lateral upper arm Intramuscular injection site on the upper arm</p>	2	pc		

	<p>Arterial system including the radial and brachial arteries</p> <p>Suture and incision sites on both the upper arm and forearm</p> <p>Hemodialysis site on the forearm</p> <p>Arterial and Venous insert for IV training, venipuncture practice, blood draw exercises, AV anastomosis, and placement of AV grafts. This multi-layer surgical insert includes the skin, subcutaneous tissue, muscle, radial artery, and radial vein</p> <p>AV fistula insert that simulates a healed fistula for hemodialysis exercises. Multi-layer bicep insert that includes the skin, subcutaneous tissue, and muscle and allows incision practice and suturing exercises.</p> <p>Durable skin that can be pierced in excess of 200 times with a 20 or 22 gauge needle.</p> <p>Realistic tactile feedback for both surgical and arterial & venous stick exercises.</p> <p>Adjustable heart rate and pulse strength simulating a heart rate from 10 BPM to 150 BPM.</p> <p>Cephalic (antecubital), Basilic, Radial, and Ulnar veins, as well as the radial and brachial arteries for infusion and blood draw training. Rotating arm allowing dorsal and volar access along the length of the IV training arm. Varying vessel palpability to simulate collapsed or bulging vessels. Ease of assembly. Latex-Free New proprietary materials replicate the skin, subcutaneous, and muscle layers at all surgical sites.</p> <p>Upgraded inserts, skins, and vessels for improved tactile feedback</p> <p>Micropump embedded within the shoulder generates variable heart rates and pulse strength. Quiet and compact, thereby improving the portability of the unit.</p> <p>Easy to replace plug-and-play inserts to allow quick change-out between procedures. Latex-free vessels with improved access for hassle-free replacement</p>				
7.	<p>CPR Manikins (set of 3 manikins-Adult, Child & Infant)</p> <p>Features:</p> <ul style="list-style-type: none"> • CPR feedback monitor that meets 2015 AHA CPR guidelines for Adult Manikin. • Available in Adult, Child, and Infant models. • Realistic looking and performing manikin is more acceptable to both instructors and students. • Far easier and faster to carry and set up at training sites. 	3	pc		

<p>•Patented face/head tilt simulates how an actual victim’s head would move if he required CPR and teaches students how to properly open the airway.</p> <p>•Easily inserted face-shield/lung-bag provides visual chest rise to ensure students receive the most accurate and realistic training.The face shield is secured so it won’t slip and there’s no need for adhesives.Prestan Manikins are not made with natural rubber latex</p> <p>2 Extra Compression Springs (Hard and Soft) Training Pads, Airway Lubricant, Eye Set, Kneel Mat, Jacket, Pants Manikin Wipes, USB Cable Plus a ShockLink with Defibrillator Training Pads for ShockLink</p> <p>ShockLink Remote Control</p> <p>ShockLink Pouch and User Guide</p> <p>Includes</p> <p>1 x Adult Manikin with CPR monitor</p> <p>1 x Child Manikin with CPR monitor</p> <p>1 x Infant Manikin with CPR monitor</p> <p>10 x Adult Face Shield Lung Bags</p> <p>10 x Child Face Shield Lung Bags</p> <p>10 x Infant Face Shield Lung Bags</p> <p>1 x Nylon Carrying case</p> <p>Warranty: 1 year</p>				
Total Bid Price for Lot 1				

[Signature]
[Name of Authorized Signatory]
[Position/Title of Authorized Signatory]

[Bidder's Letterhead]

[Date]

To: Tarlac State University
Re: Invitation to Bid No.

SCHEDULE OF PRICES

Item No.	Description	Qty.	Unit	Unit Price	Total Price
8.	<p>BABY BASSINET with stand single Bassinet stand with acrylic glass basket Bassinet stand dimension: 83 x 46 x 64.5cm Frame leg and support stainless steel Shelf-stainless steel Mattress of the bassinet is ordinary foam with cloth cover With four lockable caster wheels at 4 inches in size Basket dimension: (H) 22cm x (L) 75.5cm x (W) 40cm Warranty: 1 year</p>	4	pc		
9.	<p>EMERGENCY BED STRECHER Operated by Handle. When you turn it right, it is up, turn left, it is down. The Central-Lock System and Swivel Castor can help the User easy to control and stop. Wheel Size: Φ150mm Double Sided. High position: 192x61x83cm Low position: 192x63x54cm Size of the castors: Dia.150mm Max angle of the back: 65° Self-weight: 85kgs Load bearing: 159kgs Packing Quantity: 1 Size(LxWxH): 198x73x64cm Warranty: 1 year</p>	2	pc		
10.	<p>Medical Emergency Crash Cart 1. Mainly of aluminum, steel and ABS engineering plastic; bear Four columns to bear weight; 2. Upper part: ABS integrated board molding by injection one-time, with handralls on both sides, small items won't slip; With transparent soft glass on the table effectively prevents scratches 3. Front: Push and pull smoothly, the Centralized Lock, Key can be folded to ensure the safety of the items inside the drawer; The drawer handle is dovetail shaped, which can effectively prevent dust from entering. Five drawers, which are</p>	3	pc		

	<p>divided into two small, two medium, and one large drawers. Each drawer inside with 3*3 dividers, can be separated freely. Label card holder is slot type design, to prevent liquid and dust into holder. The label card holder is inclined upward for convenient looking, which is according to the ergonomic principle.</p> <p>4. Left side: Defibrillator shelf, which is used to store the defibrillator. It has fixed magic straps and does not fall off. Pull-out sliding writing frame improves space application efficiency Movable File Container enables easier access to patient information</p> <p>5. Right side: Equipped with lifting infusion stand, convenient and easy to operate; One 2L needle disposal holder in the utility basket, Double ABS dust basket for waste sorting will effectively eliminate the smell of garbage;</p> <p>6. Back:</p> <ul style="list-style-type: none"> -CPR Board, for acute stroke patients - Oxygen Tank Holder with belt attached to the body of the trolley. Does not drop when moving. -power outlet & hook with 5m wire. <p>7. Bottom:</p> <ul style="list-style-type: none"> -luxury universal insert noiseless casters, two with brake, casters are made of high-strength polyurethane. -Anti- static, anti-hair winding, portable and flexible; 750*475*930mm 5 Layers Packing Size: 760*540*920 mm Rack of multi bin container <p>Disposable lock with apple-shape, IV Pole, Defibrillator shelf</p> <p>Warranty: 1 year</p>				
11.	<p>INFUSION PUMP (1) and SYRINGE PUMP(1)</p> <p>Extendable from 3 up to 6 or 9 Pump Slots</p> <ul style="list-style-type: none"> • by simply connecting 2 or 3 Standard Racks (TE*RS700N) or by adding 1 or 2 Rack Extension Devices to the Communication Rack <p>Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of pump & sets* Stable infusion thanks to dedicated set</p> <p>Included accessories AC power cable Pole clamp (screw type) Instruction manual</p> <p>Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877</p> <p>Luer-Lock syringes Infusion Sets Included accessories AC power cable Pole clamp</p>	2	pc		

	<p>(screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of Terumo pump & sets* Stable infusion thanks to dedicated set * Flow rate accuracy: within $\pm 5\%$ Available in different sizes Transparent barrel Bold and easy legible scale markings No Latex & PVC components Silicon coating Double contact gasket</p> <p>Infusion Sets Included accessories AC power cable Pole clamp (screw type) Instruction manual Optional accessories Drip sensor TE*977 (standard for LF series) One touch pole clamp TE*877 Easy setting following the instruction screen of the pump Adjustable roller clamp with notch to mount tubing during priming AFF clip for prevention of accidental bolus infusion Reliable high accuracy combination of pump & sets* Stable infusion t* Flow rate accuracy: within $\pm 5\%$ Available in different sizes Transparent barrel Bold and easy legible scale markings No Latex & PVC components Silicon coating Double contact gasket</p> <p>Warranty: 1 year</p>				
12.	<p>INFANT DIGITAL WEIGHING SCALE</p> <p>Special feature Backlit Display Material: Plastic Item weight: 8 Pounds Small Footprint. Accuracy to 1 2 ounce. Detachable tray for easy cleaning and portability. Ergonomically designed for both ease of use and infant safety. features: metric lockout, integral measuring tape (up to 22 inch/56 cm), 30 lb x 0.1 oz / 15 kg x 0.005 kg capacity, locking pin for easily removing the tray for cleaning, 1 inch (25 mm) high LCD, and RS232 serial output for connectivity with a printer or PC. The spacious weighing tray measures 26 in W x 12 in D / 66 cm W x 31 cm D. The scale's ZERO button removes blanket tare weights. Your choice of rechargeable NiCad batteries (not included), standard Alkaline batteries (not included), or UL-listed AC adapter (included). The 6745 pairs perfectly with the Detecto stainless steel baby scale cart.</p> <p>Warranty: 1 year</p>	2	pc		

13.	<p>12 LEAD ECG MACHINE Channels Record & print 12 channels of data at a time, fastest speed Standard 12 leads; Cabrera; Nehb, VCG 9 Inches, 15° angled display for clear visualization Touchable, time-saving operation quick input of patient data Rechargeable, 3-hour continuous work Local Storage: 3,500 files Thermal Paper:Z-fold paper: 210*140mm Economic Recording Mode Weight: 3.4KG Accessories in the Package ECG Acquisition Box with Lead Wire Z-fold Thermal Paper (160 Pieces: 112mm × 140mm) Power Cord ×2 (US & EU Standard) & AC Power Adapter Equipotential Grounding Cable Warranty: 1 year</p>	2	pc		
14.	<p>Orthopedic bed with balkan frame, 3 cranks ·Firm mattress ·Fracture board ·Bed elevator/ shock block (optional) ·Balkan frame (cross bar, curve bar, vertical, horizontal, diagonal bars) ·3 pulleys ·3 clamps ·Overhead trapeze Dimensions: L2240xW1060xH470=680mm Function : 3 functions Material : Cold rolled steel, HDPE, Aluminum alloy Side rails : ABS Side Rails Material: Cold rolled steel, HDPE, Aluminum alloy Dimensions : L2240xW1060xH470=680mm Bed platform : 1980x900mm Backrest : 0-75deg +-5deg Legrest : 0-45deg +-5deg Height adjustment : 470-680+-10mm Brake system : Central brake Sticker color : Wood Grain/Lake Blue Side rails : ABS side rails Orthopedics traction shelf : Aluminum Alloy 2023”793”1503mm Features 1. Blue liquid angle indicator on backrest and leggiest side rails 2. Clear acrylic head card for patient name display, four infusion holes, four drainage hooks around the bed 3. Mechanical crank operation with overload protection 4. The screw handle is foldable under the</p>	1	set		

	bed footboard 5. Four 125mm diameter castors with central brake 6. Orthopedics traction is made of molding aluminum alloy 2023*793*1503mm Warranty: 1 year				
15.	Balance Skeletal Traction Equipment ·Thomas splint (with half ring) ·Pearson's attachment ·Rest splint ·Steinmann's pin holder ·Braun-Bohler splint ·Slings with clips or pins ·3 sash cords (thigh rope, traction rope, suspension rope) ·2 weight bags Traction weight (10% of patient's body weight) Suspension weight (1/2 of traction weight) ·Foot pedal or foot board Warranty: 1 year	2	set		
16.	Limb (Foot) use for Balance Skeletal Traction (1 adult and 1 Pedia Limb Warranty: 1 year	2	pc		
17.	OR LIGHT SINGLE HEAD max illumination power 160,000lux diameter of lighthouse 670mm color temperature 3000k-5000k color rendering index 96% dimming range 1-100% light field diameter d'10 24-36cm/9.5"-14.2" light field diameter d50 12-28cm/5.1"-11" number of leds 60leds lifetime of led >60,000hours degree of turbulence (standard value 37.5%max) 29% depth of volume of light L1 + L2 at 20% 130cm/51.2" depth of volume of light L1 + L2 at 60% 76cm/30" irradiance at nominal illuminance 510w/m2 handle holder made of ABS plastic flexible arms that provide comfortable positioning there is a laser marking feature to mark the midpoint of the central illumination intensity endo light mode/standby mode can be switch by a single button Warranty: 1 year	1	pc		
18.	OPERATING ROOM TABLE Electro Hydraulic OR Table Easy to handle modular design for different surgical disciplines Long and strong rail line for accessories	1	pc		

	<p>T base design for maximum approach Hand control and column control units Battery back up: up to 30 full movements</p> <p>Technical Features: Sliding Movement: 300mm Lateral Tilt Movement: +20 degree Celsius / -20 degree Celsius Back Rest Movement: +80 degree Celsius / -45 degree Celsius Trendelenburg: +30 degree Celsius / -80 degree Celsius Reverse Trendelenburg: -30 degree Celsius / -90 degree Celsius Leg Rest Up/Down: +30 degree Celsius / -80 degree Celsius Head Rest Movement: +30 degree Celsius / -90 degree Celsius Flex/Reflex Position: +225 degree Celsius / -100 degree Celsius Width x Length: 560mm x 2200mm Height (Min): 595mm Height (Max): 1035mm Mattress Thickness: 70mm Castor Diameter: 100mm Maximum Patient Weight: 360kg Stainless Steel T Base Geopel Leg Holder -made of high-quality stainless steel -pair, including clamp -height adjustable -rotatable and tillable via ball-join Warranty: 1 year</p>				
19.	<p>Digital Binocular Biological Microscope Optical System: Infinite Optical System Eyepiece: WF10x / 20mm, ANTI FUNGUS with pointer Viewing Head: Seidentopf Binocular Head, inclined at 30°, Interpupillary 47-78mm ANTI FUNGUS Objective: Infinite Semi-Plan Achromatic Objectives 4x, 10x, 40x, 100x ANTI FUNGUS Nosepiece: Backward Quadruple Nosepiece Stage: RACKLESS Double Layers Mechanical Stage 150mm×139mm, Moving Range 75mm×52mm Condenser: Abbe Condenser NA1.25 Focusing: Coaxial Coarse and Fine Adjustment, Fine Division 20mm, Moving Range 0.002mm Illumination: S-LED Illumination with Rechargeable Battery, Brightness Adjustable, can be connected to an LCD monitor or HD TV via HDMI cable and operated</p>	2	pc		

	<p>independently without connecting to PC. The image/video capture and operate can be controlled by mouse, so no shaking when you take images and videos. It also can be connected to a PC via USB2.0 cable and operate with the software.</p> <p>Features</p> <p>Use mouse to control the Camera.</p> <p>Record image and video to SD card.</p> <p>High frame rate of 15fps.</p> <p>The functions inside camera (Cloud 1.0)The less Icons the Better.</p> <p>Set Exposure Time Capability.</p> <p>3D Noise Reduction.</p> <p>1080P Video Recording.</p> <p>Get more Details with ROI Magnification Function.</p> <p>Image Comparison Function.</p> <p>Browse Captured Images</p> <p>PC software</p> <p>Warranty: 1 year</p>				
20.	<p>Electric Wheelchair</p> <p>Total length of wheelchair: 920MM</p> <p>Total height of wheelchair body: 850MM</p> <p>Seat width: 460MM</p> <p>Rear wheel size: 12 inch</p> <p>Front wheel size: 8 inch</p> <p>Frame material: steel</p> <p>Motor power: 250W (2) 60HZ</p> <p>Climbing capacity: 35 °</p> <p>Net weight 30KG</p> <p>Both sides open and close armrest</p> <p>Breathable cushion</p> <p>High-strength ABS anti-lashing rear wheel</p> <p>Thickened footrest</p> <p>Thickened steel</p> <p>Adjustable seat belt</p> <p>Back-up reminder</p> <p>Large capacity back pocket</p> <p>Anti-dumping rear wheel</p> <p>EPBS braking system without skidding</p> <p>Warranty: 1 year</p>	2	pc		
21.	<p>WEIGHING SCALE WTH HEIGHT MEASUREMENT</p> <p>Heavy-duty solid stable platform</p> <p>Dual-reading die-cast weight beam which may be read from either side of the scale</p> <p>Durable steel construction with an electrostatic powder paint finish for optimum quality.</p> <p>Capacity may be increased by using an optional counterweight</p> <p>Capacity: 400 lb x 4 oz , 180 kg x 100 g, 400 lb x 4 oz / 175 kg x 100 g</p> <p>Scale Height: 59 in / 150 cm</p>	1	pc		

	<p>Height Rod (on select models - 2392 excluded): 30 to 78 in / 76 to 200 cm - 1/4 in / 1/2 cm increments</p> <p>Platform Size: 10.5 in W x 14.5 in D / 27 cm W x 37 cm</p> <p>Warranty: 1 year</p>				
22.	<p>HOSPITAL BED</p> <p>Electric multi-functional nursing bed medical bed hospital bed turn over bed rehabilitation lift bed</p> <p>Intelligent remote control operation (also supports manual)</p> <p>Sending shampoo basin, universal wheel, urinal, infusion rack, dining board</p> <p>International certified motor (38DB) low noise</p> <p>Can be timed to turn over (30-45 minutes once)</p> <p>Electric start back</p> <p>One piece reset</p> <p>Electric lift leg (0-25°) Electric drop leg (0-85°)</p> <p>Electric turning (0-45°)</p> <p>Electric helper for defecation</p> <p>Head washing function</p> <p>Foot washing function</p> <p>Infusion function</p> <p>Meal function</p> <p>Mobility function</p> <p>Automatic lifting of the guardrail</p> <p>Anti-wetting mattress</p> <p>Size:2060*1030*480mm Bed size:1950*900mm.</p> <p>Warranty: 1 year</p>	4	pc		
23.	<p>36" Discover Headwall Package</p> <p>PRODUCT FEATURES</p> <ul style="list-style-type: none"> •The Functioning Wall Mounted 36" Discover Headwall Package is the most popular training headwall. It simulates oxygen, air and suction with one source of air by connecting to an external Silent Air Compressor. The Discover headwall package is designed for simulation labs so it is easy to install and maintain, and is available in a variety of sizes and color options, Discover size: 39.38" H x 19.5" W x 4.75" D, Wall mounted: Mounting strip included, Nurse call facade with code blue button and hand pendant <p>TECHNICAL DATA</p> <ul style="list-style-type: none"> • Size: 36", Wall Mounted – Includes mounting strip, Ohmeda gas outlets, Oxygen, Air, Vacuum, Duplex receptacle, Vacuum slide, Nurse call 3 gang cover, Includes Complete Accessories <p>Package</p> <ul style="list-style-type: none"> •Oxygen Flow Meter, Medical Air Flow 	1	package		

	Meter, Analog Vacuum Regulator, Suction Canister, Humidifiers, Nasal Cannulas, Suction tubing, DISS Hose and Fittings Warranty: 1 year				
Total Bid Price for Lot 2					

[Signature]
 [Name of Authorized Signatory]
[Position/Title of Authorized Signatory]